



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108443

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	03/25/2020	16:05
2	CAL2	1	Water	NA	NA	20.2	7.00	03/25/2020	16:06
3	CAL3	1	Water	NA	NA	20.2	10.03	03/25/2020	16:07
4	ICV	1	Water	NA	NA	20.2	7.01	03/25/2020	16:10
5	CCV1	1	Water	NA	NA	20.2	2.01	03/25/2020	16:11
6	L2031-07	1	Water	NA	NA	19.3	7.28	03/25/2020	16:12
7	L2031-07DUP	1	Water	NA	NA	19.4	7.29	03/25/2020	16:13
8	L2031-08	1	Water	NA	NA	19.5	7.59	03/25/2020	16:15
9	L2031-09	1	Water	NA	NA	19.5	7.71	03/25/2020	16:16
10	L2031-10	1	Water	NA	NA	19.5	7.59	03/25/2020	16:19
11	L2031-11	1	Water	NA	NA	18.7	7.13	03/25/2020	16:20
12	CCV2	1	Water	NA	NA	20.2	12.01	03/25/2020	16:22

18108643

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH WL2031

WorkList ID : 137356

Department : Wet-Chemistry

Date : 03-25-2020 15:57:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/25/2020	Water	L2031-07	pH	Cool 4 deg C	PSEG01	F51	30947	03/25/2020	9040C
03/25/2020	Water	L2031-08	pH	Cool 4 deg C	PSEG01	F51	30948	03/25/2020	9040C
03/25/2020	Water	L2031-09	pH	Cool 4 deg C	PSEG01	F51	30949	03/25/2020	9040C
03/25/2020	Water	L2031-10	pH	Cool 4 deg C	PSEG01	F51	30950	03/25/2020	9040C
03/25/2020	Water	L2031-11	pH	Cool 4 deg C	PSEG01	F51	30927	03/25/2020	9040C

Date/Time 03/25/2020 16:00

Received by: 

Relinquished by: 

Date/Time 03/25/2020 18:20

Received by: 

Relinquished by: 