



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108450

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/26/2020	09:20
2	CAL2	1	Water	NA	NA	20.2	7.00	03/26/2020	09:22
3	CAL3	1	Water	NA	NA	20.2	10.03	03/26/2020	09:24
4	ICV	1	Water	NA	NA	20.1	7.01	03/26/2020	09:25
5	CCV1	1	Water	NA	NA	20.2	2.01	03/26/2020	09:26
6	L2042-10	1	Water	NA	NA	16.1	6.59	03/26/2020	09:28
7	L2042-10DUP	1	Water	NA	NA	16.2	6.60	03/26/2020	09:29
8	L2042-11	1	Water	NA	NA	15.8	6.74	03/26/2020	09:30
9	CCV2	1	Water	NA	NA	20.2	12.02	03/26/2020	09:33