



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: JIGNESHDate: 03/27/2020OVEN IN TEMP1: 98 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 03/26/2020 16:00 DATE & TIME OUT1: 03/27/2020 07:15OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 03/27/2020 08:00 DATE & TIME OUT2: 03/27/2020 09:30Run Number: LB108454BalanceID: WC SC-3OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB108454BL	LB108454BL	78.5697	78.5697	100	78.5697	78.5697	78.5697	0.0000	0
2	L2048-02	GW-1- (DISCHARGE)	74.8035	74.8035	100	74.8219	74.8219	74.8219	0.0184	184
3	L2048-02DUP	GW-1- (DISCHARGE) DUP	84.7724	84.7724	100	84.7910	84.7910	84.7910	0.0186	186

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

13108454

WorkList Name : TS L2048 WorkList ID : 137382 Department : Wet-Chemistry Date : 03-26-2020 12:22:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/01/2020	Water	L2048-02	TS	Cool 4 deg C	DVIR01	F62	GW-1-(DISCHARGE)	03/25/2020	SM2540 B

Date/Time 03/26/2020 12:22:40
 Received by: JH
 Relinquished by: JH

Date/Time 03/26/2020 16:00
 Received by: JH
 Relinquished by: JH