


7470A

LB108455
INSTRUMENT ID: CV1

Linear ▼



A= 0.0000e+000

B= 2.5691e-004

C= 2.8904e-002

Rho= 0.9998847

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.039	0.039	40	0.000	40				
0.2	0.200	0.233	0.033	794	0.0 %	794				
2.5	2.500	2.453	-0.047	9434	0.0 %	9434				
5.0	5.000	4.972	-0.028	19240	0.0 %	19240				
7.5	7.500	7.421	-0.079	28772	0.0 %	28772				
10.0	10.000	10.083	0.083	39135	0.0 %	39135				

LB108455 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Mar 2020 14:04:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	40	-	0.0000	7470A	PPB	26 Mar 2020 14:39:06	S	Std
0.2 - 1	SD.2	794	-	0.2000	7470A	PPB	26 Mar 2020 14:41:14	S	Std
2.5 - 1	52.5	9434	-	2.5000	7470A	PPB	26 Mar 2020 14:43:21	S	Std
5.0 - 1	55	19240	-	5.0000	7470A	PPB	26 Mar 2020 14:45:28	S	Std
7.5 - 1	57.5	28772	-	7.5000	7470A	PPB	26 Mar 2020 14:47:35	S	Std
10.0 - 1	58.5	39135	-	10.0000	7470A	PPB	26 Mar 2020 14:49:42	S	Std
ICV78 - 1	ICV78	15019	3.8874	-7470A	PPB	26 Mar 2020 14:52:25	U	SMPL	
ICB78 - 1	ICB78	-183	-0.0181	-7470A	PPB	26 Mar 2020 14:54:30	U	SMPL	
CCV05 - 1	CCV05	18891	4.8821	-7470A	PPB	26 Mar 2020 14:56:36	U	SMPL	
CCB05 - 1	CCB05	-173	-0.0155	-7470A	PPB	26 Mar 2020 14:58:42	U	SMPL	
CRA - 1	CRA	742	0.2195	-7470A	PPB	26 Mar 2020 15:00:48	U	SMPL	
HighStd - 1	HighStd	38288	9.8654	-7470A	PPB	26 Mar 2020 15:02:53	U	SMPL	
ChkStd - 1	ChkStd	25620	6.6109	-7470A	PPB	26 Mar 2020 15:04:59	U	SMPL	
PB127836BL - 1	PBW	-254	-0.0364	-7470A	PPB	26 Mar 2020 15:07:07	U	SMPL	
PB127836BS - 1	LCSW	14878	3.8512	-7470A	PPB	26 Mar 2020 15:09:14	U	SMPL	
L1753-16 - 1	CL-02-032520	-165	-0.0135	-7470A	PPB	26 Mar 2020 15:11:21	U	SMPL	
L1753-16DUP - 1	CL-02-032520DUP	-32	0.0207	-7470A	PPB	26 Mar 2020 15:13:28	U	SMPL	
L1753-16MS - 1	CL-02-032520MS	13966	3.6169	-7470A	PPB	26 Mar 2020 15:15:36	U	SMPL	
L1753-16MSD - 1	CL-02-032520MSD	13794	3.5727	-7470A	PPB	26 Mar 2020 15:17:44	U	SMPL	
L1761-15 - 1	SU-02-032520	-157	-0.0114	-7470A	PPB	26 Mar 2020 15:19:49	U	SMPL	
CCV06 - 1	CCV06	19172	4.9543	-7470A	PPB	26 Mar 2020 15:21:55	U	SMPL	
CCB06 - 1	CCB06	-191	-0.0202	-7470A	PPB	26 Mar 2020 15:24:01	U	SMPL	
L2031-18 - 1	DEBRIS-COMP	-45	0.0173	-7470A	PPB	26 Mar 2020 15:26:08	U	SMPL	
L2031-20 - 1	SOIL-COMP	-13	0.0256	-7470A	PPB	26 Mar 2020 15:28:14	U	SMPL	
L2033-01 - 1	WATER-RECLAMATION-TANKS	278	0.1003	-7470A	PPB	26 Mar 2020 15:30:21	U	SMPL	
L2033-02 - 1	OIL-WATER-SEPARATOR	538	0.1671	-7470A	PPB	26 Mar 2020 15:32:27	U	SMPL	
L2033-03 - 1	SLUDGE-PIT	182	0.0757	-7470A	PPB	26 Mar 2020 15:34:34	U	SMPL	
L2033-04 - 1	ZIPPER-DRAINS	17	0.0333	-7470A	PPB	26 Mar 2020 15:36:41	U	SMPL	
L2033-05 - 1	BUS-WASH	65	0.0456	-7470A	PPB	26 Mar 2020 15:38:48	U	SMPL	
L2040-03 - 1	RO-350	17	0.0333	-7470A	PPB	26 Mar 2020 15:40:56	U	SMPL	
L2040-06 - 1	RO-06	32	0.0371	-7470A	PPB	26 Mar 2020 15:43:04	U	SMPL	
L2040-09 - 1	RO-11986	22	0.0346	-7470A	PPB	26 Mar 2020 15:45:10	U	SMPL	
CCV07 - 1	CCV07	19160	4.9512	-7470A	PPB	26 Mar 2020 15:47:16	U	SMPL	
CCB07 - 1	CCB07	-47	0.0168	-7470A	PPB	26 Mar 2020 15:49:22	U	SMPL	
L2040-12 - 1	RO-346	5	0.0302	-7470A	PPB	26 Mar 2020 15:51:28	U	SMPL	
L2040-15 - 1	RO-209	17	0.0333	-7470A	PPB	26 Mar 2020 15:53:35	U	SMPL	
L2042-05 - 1	SB-01-1.0-1.5	33	0.0374	-7470A	PPB	26 Mar 2020 15:55:41	U	SMPL	
L2042-06 - 1	SB-02-1.0-1.5	28	0.0361	-7470A	PPB	26 Mar 2020 15:57:47	U	SMPL	
L2042-07 - 1	SB-03-5.5-6.0	19	0.0338	-7470A	PPB	26 Mar 2020 15:59:54	U	SMPL	
L2042-08 - 1	SB-04-5.5-6.0	29	0.0364	-7470A	PPB	26 Mar 2020 16:02:01	U	SMPL	
L2043-04 - 1	TP-2A	32	0.0371	-7470A	PPB	26 Mar 2020 16:04:08	U	SMPL	
L2043-08 - 1	TP-1A	20	0.0340	-7470A	PPB	26 Mar 2020 16:06:16	U	SMPL	
CCV08 - 1	CCV08	18857	4.8734	-7470A	PPB	26 Mar 2020 16:08:24	U	SMPL	
CCB08 - 1	CCB08	-87	0.0066	-7470A	PPB	26 Mar 2020 16:10:30	U	SMPL	
L1753-16LX5 - 1		0	0.0289	-7470A	PPB	26 Mar 2020 16:12:36	U	SMPL	
L1753-16A - 1		15085	3.9043	-7470A	PPB	26 Mar 2020 16:14:42	U	SMPL	
PB127836TB - 1	PB127813TB	-33	0.0204	-7470A	PPB	26 Mar 2020 16:16:48	U	SMPL	
CCV09 - 1	CCV09	19266	4.9785	-7470A	PPB	26 Mar 2020 16:18:55	U	SMPL	
CCB09 - 1	CCB09	-122	-0.0024	-7470A	PPB	26 Mar 2020 16:21:01	U	SMPL	