



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108459

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/26/2020	14:45
2	CAL2	1	Water	NA	NA	20.1	7.00	03/26/2020	14:46
3	CAL3	1	Water	NA	NA	20.2	10.03	03/26/2020	14:47
4	ICV	1	Water	NA	NA	20.2	7.00	03/26/2020	14:48
5	CCV1	1	Water	NA	NA	20.2	2.01	03/26/2020	14:50
6	L2050-01	1	Water	NA	NA	18.9	6.93	03/26/2020	14:52
7	L2050-01DUP	1	Water	NA	NA	18.8	6.94	03/26/2020	14:53
8	L2050-02	1	Water	NA	NA	19.7	6.98	03/26/2020	15:00
9	L2050-03	1	Water	NA	NA	19.2	7.11	03/26/2020	15:02
10	L2050-04	1	Water	NA	NA	20.0	6.49	03/26/2020	15:05
11	L2050-05	1	Water	NA	NA	20.0	6.82	03/26/2020	15:06
12	CCV2	1	Water	NA	NA	20.2	12.02	03/26/2020	15:10

WORKLIST(Hardcopy Internal Chain)

2B 108459

WorkList Name : PH L2050 WorkList ID : 137387 Department : Wet-Chemistry Date : 03-26-2020 14:34:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2020	Water	L2050-01	✓ pH	Cool 4 deg C	PSEG01	F51	40359	03/26/2020	9040C
03/26/2020	Water	L2050-02	✓ pH	Cool 4 deg C	PSEG01	F51	40768	03/26/2020	9040C
03/26/2020	Water	L2050-03	✓ pH	Cool 4 deg C	PSEG01	F51	032520-002	03/26/2020	9040C
03/26/2020	Water	L2050-04	✓ pH	Cool 4 deg C	PSEG01	F51	032520-003	03/26/2020	9040C
03/26/2020	Water	L2050-05	✓ pH	Cool 4 deg C	PSEG01	F51	032520-004	03/26/2020	9040C

Date/Time 03/26/2020 14:40
 Received by: JN
 Relinquished by: JN

Date/Time 03/26/2020 16:00
 Received by: JN
 Relinquished by: JN