



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108460

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/26/2020	14:45
2	CAL2	1	Water	NA	NA	20.1	7.00	03/26/2020	14:46
3	CAL3	1	Water	NA	NA	20.2	10.03	03/26/2020	14:47
4	ICV	1	Water	NA	NA	20.2	7.00	03/26/2020	14:48
5	CCV1	1	Water	NA	NA	20.2	2.02	03/26/2020	15:20
6	L2050-07	1	Solid	20.03	20	21.2	6.44	03/26/2020	15:22
7	L2050-07DUP	1	Solid	20.04	20	21.3	6.45	03/26/2020	15:23
8	CCV2	1	Water	NA	NA	20.2	12.02	03/26/2020	15:25

LP 108460

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L2050

WorkList ID : 137389

Department : Wet-Chemistry

Date : 03-26-2020 15:06:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/26/2020	Solid	L2050-07-15	pH	Cool 4 deg C	PSEG01	F51	OILY-DEBRIS	03/26/2020	9045D

Date/Time 03/26/2020 15:10
 Received by: Jn
 Relinquished by: Jn

Date/Time 03-26-2020 16:10
 Received by: Jn
 Relinquished by: Jn