



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108470

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/26/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	03/26/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	03/26/2020	17:18
4	ICV	1	Water	NA	NA	20.2	7.00	03/26/2020	17:20
5	CCV1	1	Water	NA	NA	20.2	2.01	03/26/2020	17:22
6	L2008-01	1	Water	NA	NA	14.6	7.21	03/26/2020	17:25
7	L2008-03	1	Water	NA	NA	14.7	7.53	03/26/2020	17:30
8	L2053-09	1	Water	NA	NA	19.6	7.17	03/26/2020	17:32
9	L2053-09DUP	1	Water	NA	NA	19.7	7.18	03/26/2020	17:34
10	CCV2	1	Water	NA	NA	20.2	12.02	03/26/2020	17:35

20108470

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2053 WorkList ID : 137392 Department : Wet-Chemistry Date : 03-26-2020 17:06:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/20/2020	Water	L2008-01-D	pH	Cool 4 deg C	TULL02	F12	MH-726	03/20/2020	9040C
03/20/2020	Water	L2008-03-D	pH	Cool 4 deg C	TULL02	F12	MH-4	03/20/2020	9040C
03/26/2020	Water	L2053-09-D	pH	Cool 4 deg C	PSEG01	F52	GWTP-15	03/26/2020	9040C

Date/Time 03/26/2020 17:10
Received by: Jo
Relinquished by: Jo

Date/Time 03/26/2020 18:10
Received by: Jo
Relinquished by: Jo