


LB108473

7471B

INSTRUMENT ID: CV2

 μ Abs.:

52777



Linear

A= 0.0000e+000

B= 1.8846e-004

C= 4.1307e-002

Rho= 0.9998280

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.044	0.044	13	0.000	13				
0.2	0.200	0.226	0.026	981	0.0 %	981				
2.5	2.500	2.359	-0.141	12298	0.0 %	12298				
5.0	5.000	5.073	0.073	26700	0.0 %	26700				
7.5	7.500	7.510	0.010	39632	0.0 %	39632				
10.0	10.000	9.988	-0.012	52777	0.0 %	52777				

LB108473 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Mar 2020 09:35:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	13	-	0.0000	7471B	PPB	27 Mar 2020 09:43:58	S	Std
0.2 - 1	50.2	981	-	0.2000	7471B	PPB	27 Mar 2020 09:46:15	S	Std
2.5 - 1	52.5	12298	-	2.5000	7471B	PPB	27 Mar 2020 09:48:32	S	Std
5.0 - 1	55	26700	-	5.0000	7471B	PPB	27 Mar 2020 09:50:48	S	Std
7.5 - 1	57.5	39632	-	7.5000	7471B	PPB	27 Mar 2020 09:53:05	S	Std
10.0 - 1	510	52777	-	10.0000	7471B	PPB	27 Mar 2020 09:55:22	S	Std
ICV79 - 1	ICV79	20795	3.9603	-	7471B	PPB	27 Mar 2020 09:58:12	U	SMPL
ICB79 - 1	ICB79	-211	0.0015	-	7471B	PPB	27 Mar 2020 10:00:27	U	SMPL
CCV10 - 1	CCV10	27145	5.1570	-	7471B	PPB	27 Mar 2020 10:02:43	U	SMPL
CCB10 - 1	CCB10	-225	-0.0011	-	7471B	PPB	27 Mar 2020 10:04:58	U	SMPL
CRA - 1	CRA	963	0.2228	-	7471B	PPB	27 Mar 2020 10:07:14	U	SMPL
HighStd - 1	HighStd	53485	10.1210	-	7471B	PPB	27 Mar 2020 10:09:29	U	SMPL
ChkStd - 1	ChkStd	36393	6.8999	-	7471B	PPB	27 Mar 2020 10:11:45	U	SMPL
PB127844BL - 1	PBS	-307	-0.0165	-	7471B	PPB	27 Mar 2020 10:14:03	U	SMPL
PB127844BS - 1	LCSS	18529	3.5333	-	7471B	PPB	27 Mar 2020 10:16:20	U	SMPL
L2049-01 - 1	SB-1-(1-3)	2972	0.6014	-	7471B	PPB	27 Mar 2020 10:18:37	U	SMPL
L2049-02 - 1	SB-3-(1-3)	1841	0.3883	-	7471B	PPB	27 Mar 2020 10:20:54	U	SMPL
L2049-03 - 1	SB-5-(1-3)	2118	0.4405	-	7471B	PPB	27 Mar 2020 10:23:10	U	SMPL
L2049-04 - 1	SB-4-(1-3)	14993	2.8669	-	7471B	PPB	27 Mar 2020 10:25:28	U	SMPL
L2049-05 - 1	SB-1-(10-12)	302	0.0982	-	7471B	PPB	27 Mar 2020 10:27:43	U	SMPL
CCV11 - 1	CCV11	27757	5.2724	-	7471B	PPB	27 Mar 2020 10:29:58	U	SMPL
CCB11 - 1	CCB11	-117	0.0193	-	7471B	PPB	27 Mar 2020 10:32:14	U	SMPL
L2049-06 - 1	SB-3-(10-12)	838	0.1992	-	7471B	PPB	27 Mar 2020 10:34:30	U	SMPL
L2049-07 - 1	SB-5-(10-12)	518	0.1389	-	7471B	PPB	27 Mar 2020 10:36:45	U	SMPL
L2049-08 - 1	SB-4-(10-12)	713	0.1757	-	7471B	PPB	27 Mar 2020 10:39:01	U	SMPL
L2050-07 - 1	OILY-DEBRIS	388	0.1144	-	7471B	PPB	27 Mar 2020 10:41:16	U	SMPL
L2051-03 - 1	SB-3	539	0.1429	-	7471B	PPB	27 Mar 2020 10:43:32	U	SMPL
L2051-03DUP - 1	SB-3DUP	678	0.1691	-	7471B	PPB	27 Mar 2020 10:48:25	U	SMPL
L2051-03MS - 1	SB-3MS	20888	3.9778	-	7471B	PPB	27 Mar 2020 10:50:41	U	SMPL
L2051-03MSD - 1	SB-3MSD	20678	3.9383	-	7471B	PPB	27 Mar 2020 10:52:58	U	SMPL
L2051-04 - 1	SB-4	99	0.0600	-	7471B	PPB	27 Mar 2020 10:55:15	U	SMPL
L2051-05 - 1	SB-5	209	0.0807	-	7471B	PPB	27 Mar 2020 10:57:32	U	SMPL
CCV12 - 1	CCV12	27547	5.2328	-	7471B	PPB	27 Mar 2020 10:59:47	U	SMPL
CCB12 - 1	CCB12	-155	0.0121	-	7471B	PPB	27 Mar 2020 11:02:03	U	SMPL
L2052-02 - 1	SOIL-PILE	17828	3.4012	-	7471B	PPB	27 Mar 2020 11:04:19	U	SMPL
L2053-02 - 1	TP-13-SA	22646	4.3091	-	7471B	PPB	27 Mar 2020 11:06:34	U	SMPL
L2053-06 - 1	TP-15-SA	10329	1.9879	-	7471B	PPB	27 Mar 2020 11:08:50	U	SMPL
L2051-03LX5 - 1		76	0.0556	-	7471B	PPB	27 Mar 2020 11:11:06	U	SMPL
L2051-03A - 1		21288	4.0532	-	7471B	PPB	27 Mar 2020 11:13:22	U	SMPL
CCV13 - 1	CCV13	27555	5.2343	-	7471B	PPB	27 Mar 2020 11:15:38	U	SMPL
CCB13 - 1	CCB13	-131	0.0166	-	7471B	PPB	27 Mar 2020 11:17:55	U	SMPL