



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108474

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/27/2020	09:50
2	CAL2	1	Water	NA	NA	20.1	7.01	03/27/2020	09:51
3	CAL3	1	Water	NA	NA	20.1	10.03	03/27/2020	09:53
4	ICV	1	Water	NA	NA	20.1	7.00	03/27/2020	10:00
5	CCV1	1	Water	NA	NA	20.1	2.01	03/27/2020	10:02
6	L2049-05	1	Solid	20.03	20	21.3	6.82	03/27/2020	10:05
7	L2049-05DUP	1	Solid	20.04	20	21.4	6.83	03/27/2020	10:06
8	L2053-04	1	Solid	20.03	20	21.3	8.15	03/27/2020	10:10
9	L2053-08	1	Solid	20.04	20	21.3	8.61	03/27/2020	10:11
10	CCV2	1	Water	NA	NA	20.1	12.02	03/27/2020	10:12

2B108474

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2052 WorkList ID : 137413 Department : Wet-Chemistry Date : 03-27-2020 09:32:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/25/2020	Solid	L2049-05	Corrosivity	Cool 4 deg C	DVIR01	F52	SB-1-(10-12)	03/25/2020	9045D
03/26/2020	Solid	L2053-04	Corrosivity	Cool 4 deg C	PSEG01	F52	TP-13-SA	03/26/2020	9045D
03/26/2020	Solid	L2053-08	Corrosivity	Cool 4 deg C	PSEG01	F52	TP-15-SA	03/26/2020	9045D

Date/Time 03/27/2020 9:40
 Received by: Jn
 Relinquished by: Jn

Date/Time 03/27/2020 11:00
 Received by: Jn
 Relinquished by: Jn