


LB108477

245.1

INSTRUMENT ID: CV4

 μ Abs.:

39591



A= 0.0000e+000

B= 2.5449e-004

C= 1.5837e-002

Rho= 0.9998491

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	-78	0.000	-78				
0.05	0.050	0.058	0.008	167	0.0 %	167				
0.2	0.200	0.203	0.003	737	0.0 %	737				
2.5	2.500	2.542	0.042	9928	0.0 %	9928				
5.0	5.000	4.996	-0.004	19567	0.0 %	19567				
7.5	7.500	7.363	-0.137	28869	0.0 %	28869				
10.0	10.000	10.091	0.091	39591	0.0 %	39591				

LB108477 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 27 Mar 2020 12:10:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	-78	-	0.0000245.1	PPB	27 Mar 2020 12:42:09	S	Std
0.05 - 1	50.05	167	-	0.0500245.1	PPB	27 Mar 2020 12:44:26	S	Std
0.2 - 1	50.2	737	-	0.2000245.1	PPB	27 Mar 2020 12:46:44	S	Std
2.5 - 1	52.5	9928	-	2.5000245.1	PPB	27 Mar 2020 12:49:00	S	Std
5.0 - 1	55	19567	-	5.0000245.1	PPB	27 Mar 2020 12:51:17	S	Std
7.5 - 1	57.5	28869	-	7.5000245.1	PPB	27 Mar 2020 12:53:32	S	Std
10.0 - 1	510	39591	-	10.0000245.1	PPB	27 Mar 2020 13:06:41	S	Std
ICV80 - 1	ICV80	15351	3.9226	-245.1	PPB	27 Mar 2020 13:16:20	U	SMPL
ICB80 - 1	ICB80	-110	-0.0122	-245.1	PPB	27 Mar 2020 13:18:34	U	SMPL
CCV14 - 1	CCV14	20501	5.2332	-245.1	PPB	27 Mar 2020 13:20:50	U	SMPL
CCB14 - 1	CCB14	-256	-0.0493	-245.1	PPB	27 Mar 2020 13:23:05	U	SMPL
CRA - 1	CRA	150	0.0540	-245.1	PPB	27 Mar 2020 13:28:20	U	SMPL
HighStd - 1	HighStd	38627	9.8462	-245.1	PPB	27 Mar 2020 13:30:35	U	SMPL
ChkStd - 1	ChkStd	26825	6.8426	-245.1	PPB	27 Mar 2020 13:32:50	U	SMPL
PB127846BL - 1	PBW	-245	-0.0465	-245.1	PPB	27 Mar 2020 13:35:07	U	SMPL
PB127846BS - 1	LCSW	14826	3.7890	-245.1	PPB	27 Mar 2020 13:37:25	U	SMPL
L2008-01 - 1	MH-726	-42	0.0051	-245.1	PPB	27 Mar 2020 13:39:41	U	SMPL
L2008-01DUP - 1	MH-726DUP	14	0.0194	-245.1	PPB	27 Mar 2020 13:41:57	U	SMPL
L2008-01MS - 1	MH-726MS	15008	3.8353	-245.1	PPB	27 Mar 2020 13:44:14	U	SMPL
L2008-01MSD - 1	MH-726MSD	14044	3.5899	-245.1	PPB	27 Mar 2020 13:46:30	U	SMPL
L2008-03 - 1	MH-4	211	0.0695	-245.1	PPB	27 Mar 2020 13:48:46	U	SMPL
CCV15 - 1	CCV15	19628	5.0110	-245.1	PPB	27 Mar 2020 13:51:01	U	SMPL
CCB15 - 1	CCB15	-115	-0.0134	-245.1	PPB	27 Mar 2020 13:53:17	U	SMPL
L2008-06 - 1	MH-726	-54	0.0021	-245.1	PPB	27 Mar 2020 13:55:33	U	SMPL
L2008-08 - 1	MH-4	-47	0.0039	-245.1	PPB	27 Mar 2020 13:57:47	U	SMPL
L2048-02 - 1	GW-1-(DISCHARGE)	-10	0.0133	-245.1	PPB	27 Mar 2020 14:00:03	U	SMPL
L2008-01LX5 - 1		3	0.0166	-245.1	PPB	27 Mar 2020 14:02:18	U	SMPL
L2008-01A - 1		15039	3.8432	-245.1	PPB	27 Mar 2020 14:04:34	U	SMPL
CCV16 - 1	CCV16	19753	5.0429	-245.1	PPB	27 Mar 2020 14:06:50	U	SMPL
CCB16 - 1	CCB16	-117	-0.0139	-245.1	PPB	27 Mar 2020 14:09:07	U	SMPL