

LB108485


ISM02.4

INSTRUMENT ID: CV4

Linear

 μ Abs.:

45273



A= 0.0000e+000

B= 2.2301e-004

C= -6.1040e-003

Rho= 0.9997052

Accept = Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.018	0.018	110	0.000	110					1
0.05	0.050					149					1
0.20	0.200	0.203	0.003	939	0.0 %	939					2
2.50	2.500	2.479	-0.021	11142	0.0 %	11142					1
5.00	5.000	5.087	0.087	22839	0.0 %	22839					2
7.50	7.500	7.322	-0.178	32861	0.0 %	32861					2
10.0	10.000	10.090	0.090	45273	0.0 %	45273					1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	110	-		0 ISM02.4	PPB	3/27/2020 16:15	Std
	0.2 S0.2	939	-		0.2 ISM02.4	PPB	3/27/2020 16:17	Std
	2.5 S2.5	11142	-		2.5 ISM02.4	PPB	3/27/2020 16:19	Std
	5 S5.0	22839	-		5 ISM02.4	PPB	3/27/2020 16:22	Std
	7.5 S7.5	32861	-		7.5 ISM02.4	PPB	3/27/2020 16:24	Std
	10 S10.0	45273	-		10 ISM02.4	PPB	3/27/2020 16:26	Std
ICV	ICV	16882	3.7587 -		ISM02.4	PPB	3/27/2020 16:29	SMPL
ICB	ICB	-40	-0.015 -		ISM02.4	PPB	3/27/2020 16:31	SMPL
CCV017	CCV017	22457	5.002 -		ISM02.4	PPB	3/27/2020 16:33	SMPL
CCB017	CCB017	-38	-0.0146 -		ISM02.4	PPB	3/27/2020 16:36	SMPL
PB127879BL	PBW879	67	0.0088 -		ISM02.4	PPB	3/27/2020 16:38	SMPL
L2060-01	MBF7G8	856	0.1848 -		ISM02.4	PPB	3/27/2020 16:40	SMPL
L2060-02	MBF7G8D	718	0.154 -		ISM02.4	PPB	3/27/2020 16:42	SMPL
L2060-03	MBF7G8S	5164	1.1455 -		ISM02.4	PPB	3/27/2020 16:45	SMPL
L2060-04	MBF7G9	-76	-0.0231 -		ISM02.4	PPB	3/27/2020 16:47	SMPL
CCV018	CCV018	22022	4.905 -		ISM02.4	PPB	3/27/2020 16:49	SMPL
CCB018	CCB018	-9	-0.0081 -		ISM02.4	PPB	3/27/2020 16:52	SMPL