



## Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108486

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2628         |
| buffer solution pH 7 yellow    | W2626         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2632         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 03/27/2020 | 16:10     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 03/27/2020 | 16:11     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.03       | 03/27/2020 | 16:12     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.2             | 7.02        | 03/27/2020 | 16:15     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.2             | 2.01        | 03/27/2020 | 16:16     |
| 6   | L1762-25    | 1  | Solid  | 20.03       | 20          | 22.2             | 8.53        | 03/27/2020 | 16:19     |
| 7   | L1762-25DUP | 1  | Solid  | 20.04       | 20          | 22.3             | 8.54        | 03/27/2020 | 16:20     |
| 8   | L2055-05    | 1  | Solid  | 20.02       | 20          | 22.2             | 7.83        | 03/27/2020 | 16:22     |
| 9   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.02       | 03/27/2020 | 16:25     |

23108486

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L2055

WorkList ID : 137440

Department : Wet-Chemistry

Date : 03-27-2020 15:50:21

| Due Date   | Matrix | Sample   | Test        | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method |
|------------|--------|----------|-------------|--------------|----------|------------------|-----------------|--------------|--------|
| 03/27/2020 | Solid  | L1762-25 | Corrosivity | Cool 4 deg C | PSEG05   |                  | TR-04-032720    | 03/27/2020   | 9045D  |
| 03/26/2020 | Solid  | L2055-05 | Corrosivity | Cool 4 deg C | DVIR01   | F51              | SB-8-(1-3)      | 03/26/2020   | 9045D  |

Date/Time 03/27/2020 16:00  
 Received by: JH  
 Relinquished by: JH

Date/Time 03/27/2020 17:00  
 Received by: JH  
 Relinquished by: JH