



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: Daniel

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108501

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	03/27/2020	08:00
2	L2011-04		1	8	100.00	>212.0	>212.0	03/27/2020	08:20
3	L2011-04DUP		1	9	100.00	>212.0	>212.0	03/27/2020	08:40
4	L2021-07		1	9	100.00	>212.0	>212.0	03/27/2020	09:05
5	L2021-08		1	8	100.00	>212.0	>212.0	03/27/2020	09:30
6	L2031-22		1	9	100.00	>212.0	>212.0	03/27/2020	09:52
7	L2042-10		1	8	100.00	>212.0	>212.0	03/27/2020	10:15
8	L2042-11		1	8	100.00	>212.0	>212.0	03/27/2020	10:32
9	L2048-02		1	9	100.00	>212.0	>212.0	03/27/2020	10:55
10	L2050-06		1	8	100.00	>212.0	>212.0	03/27/2020	11:25
11	L2053-11		1	9	100.00	>212.0	>212.0	03/27/2020	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB102500 LB108501 04/30/2020

Date: 03-27-2020 10:16:46

Date/Time 03/27/2020 7:30
Received by: DK
Relinquished by: SD