


LB 108504

7470A

INSTRUMENT ID : CV 1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.070	0.070	7	0.000	7				
0.2	0.200	0.237	0.037	649	0.0 %	649				
2.5	2.500	2.467	-0.033	9236	0.0 %	9236				
5.0	5.000	4.786	-0.214	18166	0.0 %	18166				
7.5	7.500	7.629	0.129	29112	0.0 %	29112				
10.0	10.000	10.011	0.011	38286	0.0 %	38286				

LB108504 IONSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 30 Mar 2020 13:20:33					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	7	-	0.0000	7470A	PPB	30 Mar 2020 13:24:47	S	Std
0.2 - 1	50.2	649	-	0.2000	7470A	PPB	30 Mar 2020 13:31:52	S	Std
2.5 - 1	52.5	9236	-	2.5000	7470A	PPB	30 Mar 2020 13:33:59	S	Std
5.0 - 1	55	18166	-	5.0000	7470A	PPB	30 Mar 2020 13:36:06	S	Std
7.5 - 1	57.5	29112	-	7.5000	7470A	PPB	30 Mar 2020 13:38:14	S	Std
10.0 - 1	57.5	38286	-	10.0000	7470A	PPB	30 Mar 2020 13:47:59	S	Std
ICV81 - 1	ICV81	14595	3.8587	-	7470A	PPB	30 Mar 2020 13:50:38	U	SMPL
ICB81 - 1	ICB81	-248	0.0041	-	7470A	PPB	30 Mar 2020 13:52:44	U	SMPL
CCV19 - 1	CCV19	18691	4.9224	-	7470A	PPB	30 Mar 2020 13:54:50	U	SMPL
CCB19 - 1	CCB19	-247	0.0043	-	7470A	PPB	30 Mar 2020 13:56:55	U	SMPL
CRA - 1	CRA	604	0.2253	-	7470A	PPB	30 Mar 2020 13:59:01	U	SMPL
HighStd - 1	HighStd	38396	10.0396	-	7470A	PPB	30 Mar 2020 14:01:07	U	SMPL
ChkStd - 1	ChkStd	26417	6.9288	-	7470A	PPB	30 Mar 2020 14:03:13	U	SMPL
PB127853BL - 1	PBW	-316	-0.0136	-	7470A	PPB	30 Mar 2020 14:05:21	U	SMPL
PB127853BS - 1	LCSW	13843	3.6634	-	7470A	PPB	30 Mar 2020 14:07:29	U	SMPL
L2031-22 - 1	WATER-COMP	-225	0.0100	-	7470A	PPB	30 Mar 2020 14:09:36	U	SMPL
L2053-11 - 1	GWTP-15	-54	0.0545	-	7470A	PPB	30 Mar 2020 14:11:43	U	SMPL
L2053-11DUP - 1	GWTP-15DUP	-31	0.0604	-	7470A	PPB	30 Mar 2020 14:13:50	U	SMPL
L2053-11MS - 1	GWTP-15MS	14010	3.7068	-	7470A	PPB	30 Mar 2020 14:15:58	U	SMPL
L2053-11MSD - 1	GWTP-15MSD	14646	3.8719	-	7470A	PPB	30 Mar 2020 14:18:04	U	SMPL
CCV20 - 1	CCV20	18761	4.9406	-	7470A	PPB	30 Mar 2020 14:20:10	U	SMPL
CCB20 - 1	CCB20	-232	0.0082	-	7470A	PPB	30 Mar 2020 14:22:16	U	SMPL
PB127854BL - 1	PBW	-88	0.0456	-	7470A	PPB	30 Mar 2020 14:24:22	U	SMPL
PB127854BS - 1	LCSW	13982	3.6995	-	7470A	PPB	30 Mar 2020 14:26:28	U	SMPL
L2048-01 - 1	GW-1	87	0.0911	-	7470A	PPB	30 Mar 2020 14:33:05	U	SMPL
L2048-01DUP - 1	GW-1DUP	40	0.0789	-	7470A	PPB	30 Mar 2020 14:35:10	U	SMPL
L2048-01MS - 1	GW-1MS	15543	4.1049	-	7470A	PPB	30 Mar 2020 14:37:17	U	SMPL
L2048-01MSD - 1	GW-1MSD	15785	4.1677	-	7470A	PPB	30 Mar 2020 14:39:24	U	SMPL
L2048-03 - 1	GW-3	2567	0.7351	-	7470A	PPB	30 Mar 2020 14:41:30	U	SMPL
L2048-04 - 1	GW-4	81	0.0895	-	7470A	PPB	30 Mar 2020 14:43:38	U	SMPL
L2050-06 - 1	WATER-COMP	40	0.0789	-	7470A	PPB	30 Mar 2020 14:45:47	U	SMPL
L2053-09 - 1	GWTP-15	846	0.2882	-	7470A	PPB	30 Mar 2020 14:47:53	U	SMPL
CCV21 - 1	CCV21	19047	5.0148	-	7470A	PPB	30 Mar 2020 14:50:00	U	SMPL
CCB21 - 1	CCB21	-94	0.0441	-	7470A	PPB	30 Mar 2020 14:52:06	U	SMPL
PB127862BL - 1	PBW	-8	0.0664	-	7470A	PPB	30 Mar 2020 14:54:12	U	SMPL
PB127862BS - 1	LCSW	15175	4.0093	-	7470A	PPB	30 Mar 2020 14:56:18	U	SMPL
L2032-02 - 1	SOIL-CUTTINGS-1	-32	0.0602	-	7470A	PPB	30 Mar 2020 14:58:25	U	SMPL
L2052-04 - 1	SOIL-PILE	71	0.0869	-	7470A	PPB	30 Mar 2020 15:00:31	U	SMPL
L2053-04 - 1	TP-13-SA	147	0.1066	-	7470A	PPB	30 Mar 2020 15:02:37	U	SMPL
L2053-08 - 1	TP-15-SA	-35	0.0594	-	7470A	PPB	30 Mar 2020 15:04:44	U	SMPL
L2053-08DUP - 1	TP-15-SADUP	-17	0.0641	-	7470A	PPB	30 Mar 2020 15:06:51	U	SMPL
L2053-08MS - 1	TP-15-SAMS	13596	3.5992	-	7470A	PPB	30 Mar 2020 15:08:58	U	SMPL
L2053-08MSD - 1	TP-15-SAMSD	14282	3.7774	-	7470A	PPB	30 Mar 2020 15:11:06	U	SMPL
CCV22 - 1	CCV22	18767	4.9421	-	7470A	PPB	30 Mar 2020 15:13:12	U	SMPL
CCB22 - 1	CCB22	-149	0.0298	-	7470A	PPB	30 Mar 2020 15:15:19	U	SMPL
L2053-11LX5 - 1		-52	0.0550	-	7470A	PPB	30 Mar 2020 15:17:26	U	SMPL
L2053-11A - 1		14534	3.8428	-	7470A	PPB	30 Mar 2020 15:19:32	U	SMPL
L2048-01LX5 - 1		-135	0.0334	-	7470A	PPB	30 Mar 2020 15:21:38	U	SMPL
L2048-01A - 1		15631	4.1277	-	7470A	PPB	30 Mar 2020 15:23:44	U	SMPL
L2053-08LX5 - 1		-105	0.0412	-	7470A	PPB	30 Mar 2020 15:25:51	U	SMPL
L2053-08A - 1		15443	4.0789	-	7470A	PPB	30 Mar 2020 15:27:57	U	SMPL
PB127853TB - 1	PB127845TB	-25	0.0620	-	7470A	PPB	30 Mar 2020 15:30:04	U	SMPL
PB127862TB - 1	PB127817TB	25	0.0750	-	7470A	PPB	30 Mar 2020 15:32:11	U	SMPL
CCV23 - 1	CCV23	18896	4.9756	-	7470A	PPB	30 Mar 2020 15:34:18	U	SMPL
CCB23 - 1	CCB23	-151	0.0293	-	7470A	PPB	30 Mar 2020 15:36:26	U	SMPL