



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108505

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/30/2020	13:30
2	CAL2	1	Water	NA	NA	20.2	7.00	03/30/2020	13:31
3	CAL3	1	Water	NA	NA	20.2	10.04	03/30/2020	13:32
4	ICV	1	Water	NA	NA	20.1	6.99	03/30/2020	13:35
5	CCV1	1	Water	NA	NA	20.2	2.02	03/30/2020	13:36
6	L2073-03	1	Water	NA	NA	16.7	6.96	03/30/2020	13:38
7	L2073-03DUP	1	Water	NA	NA	16.8	6.97	03/30/2020	13:39
8	L2073-04	1	Water	NA	NA	17.6	6.98	03/30/2020	13:40
9	L2073-05	1	Water	NA	NA	18.2	5.60	03/30/2020	13:43
10	CCV2	1	Water	NA	NA	20.2	12.02	03/30/2020	13:45