



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108507

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	03/30/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	03/30/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	03/30/2020	17:12
4	ICV	1	Water	NA	NA	20.1	7.01	03/30/2020	17:15
5	CCV1	1	Water	NA	NA	20.2	2.02	03/30/2020	17:16
6	L1752-08	1	Solid	20.03	20	21.3	8.04	03/30/2020	17:19
7	L1752-08DUP	1	Solid	20.04	20	21.4	8.05	03/30/2020	17:20
8	L1758-06	1	Solid	20.03	20	21.4	7.73	03/30/2020	17:22
9	L1759-20	1	Solid	20.04	20	21.4	10.13	03/30/2020	17:25
10	L2070-01	1	Solid	20.04	20	21.2	6.09	03/30/2020	17:26
11	L2070-02	1	Solid	20.02	20	21.3	6.21	03/30/2020	17:29
12	L2070-03	1	Solid	20.03	20	21.1	6.56	03/30/2020	17:30
13	L2070-04	1	Solid	20.04	20	21.6	7.00	03/30/2020	17:32
14	L2072-01	1	Solid	20.02	20	22.0	7.94	03/30/2020	17:35
15	L2072-02	1	Solid	20.03	20	21.9	6.75	03/30/2020	17:38
16	CCV2	1	Water	NA	NA	20.2	12.02	03/30/2020	17:39

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1752 WorkList ID : 137455 Department : Wet-Chemistry Date : 03-30-2020 12:05:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2020	Solid	L1752-08	Corrosivity	Cool 4 deg C	PSEG05		HD-02-033020	03/30/2020	9045D
03/30/2020	Solid	L1758-06	Corrosivity	Cool 4 deg C	PSEG05	F51	PL-03-033020	03/30/2020	9045D
03/30/2020	Solid	L1759-20	Corrosivity	Cool 4 deg C	PSEG05		OR-03-033020	03/30/2020	9045D
03/26/2020	Solid	L2070-01	Corrosivity	Cool 4 deg C	CLEA11	F61	BUSH-WASH	03/26/2020	9045D
03/26/2020	Solid	L2070-02	Corrosivity	Cool 4 deg C	CLEA11	F61	RECLAIM-SYSTEM-SLUDGE	03/26/2020	9045D
03/26/2020	Solid	L2070-03	Corrosivity	Cool 4 deg C	CLEA11	F61	CHASSIS-WASH	03/26/2020	9045D
03/26/2020	Solid	L2070-04	Corrosivity	Cool 4 deg C	CLEA11	F61	OIL-WATER-SEPERATOR-SL	03/26/2020	9045D
03/27/2020	Solid	L2072-01	Corrosivity	Cool 4 deg C	CLEA11	F31	MECHANIC-BUS-LIFT	03/27/2020	9045D
03/27/2020	Solid	L2072-02	Corrosivity	Cool 4 deg C	CLEA11	F31	BUS-WASH	03/27/2020	9045D

Date/Time 03/30/2020 16:00
 Received by: JN
 Relinquished by: JN

Date/Time 03/30/2020 18:10
 Received by: JN
 Relinquished by: JN