



PERCENT SOLID

Supervisor: Sohil
Analyst: JIGNESH
Date: 3/31/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 03/30/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 03/31/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108510

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1752-05	HD-02-033020	1	1.12	9.58	8.25	84.3	
L1752-06	HD-02-033020	2	1.18	9.74	7.94	79.0	
L1752-07	HD-02-033020-1	3	1.13	9.82	9.26	93.6	
L1758-01	PL-03-033020	4	1.13	9.94	8.84	87.5	
L1758-02	PL-03-033020	5	1.17	9.89	8.71	86.5	
L1758-03	PL-03-033020	6	1.15	9.65	8.46	86.0	
L1758-04	PL-03-033020	7	1.16	9.88	8.75	87.0	
L1758-05	PL-03-033020	8	1.17	9.98	9.46	94.1	
L1759-16	OR-03-033020	9	1.17	9.74	8.84	89.5	
L1759-17	OR-03-033020	10	1.15	9.94	9.26	92.3	
L1759-18	OR-03-033020-1	11	1.16	9.63	8.89	91.3	
L1759-19	OR-03-033020-2	12	1.16	9.58	8.52	87.4	
L2007-14	K205-7C	13	1.14	9.9	7.63	74.1	
L2007-16	K205-8C	14	1.13	9.57	7.85	79.6	
L2007-18	K205-9C	15	1.15	9.95	8.19	80.0	
L2007-20	K205-10C	16	1.13	9.79	7.78	76.8	
L2007-22	K205-11C	17	1.17	9.96	8.34	81.6	
L2067-01	239-SB-01-16-17	18	1.18	9.97	9.00	89.0	
L2067-02	239-SB-01-4.5-5.5	19	1.17	9.65	8.45	85.8	
L2071-01	INTERIOR-WALL	20	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L2071-03	EXTERIOR-WALL	21	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L2088-01	WASTE-1	22	1.15	9.94	8.62	85.0	
L2088-02	1	23	1.17	9.52	8.4	86.6	
L2088-03	2	24	1.15	9.69	8.5	86.1	
L2088-04	3	25	1.15	9.93	8.85	87.7	
L2088-05	4	26	1.1	9.74	8.48	85.4	
L2088-06	5	27	1.14	9.94	8.8	87.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$