


LB108511
INSTRUMENT ID: CY2

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.030	-0.030	56	0.000	56				
0.2	0.200	0.161	-0.039	1063	0.0 %	1063				
2.5	2.500	2.437	-0.063	13048	0.0 %	13048				
5.0	5.000	5.187	0.187	27527	0.0 %	27527				
7.5	7.500	7.625	0.125	40366	0.0 %	40366				
10.0	10.000	9.820	-0.180	51923	0.0 %	51923				

LB108511 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 30 Mar 2020 16:00:41					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	56	-	0.0000	7471B	PPB	30 Mar 2020 16:03:45	S	Std
0.2 - 1	50.2	1063	-	0.2000	7471B	PPB	30 Mar 2020 16:06:02	S	Std
2.5 - 1	50.5	13048	-	2.5000	7471B	PPB	30 Mar 2020 16:08:19	S	Std
5.0 - 1	55	27527	-	5.0000	7471B	PPB	30 Mar 2020 16:13:18	S	Std
7.5 - 1	57.5	40366	-	7.5000	7471B	PPB	30 Mar 2020 16:15:34	S	Std
10.0 - 1	57.5	51923	-	10.0000	7471B	PPB	30 Mar 2020 16:17:52	S	Std
ICV82 - 1	ICV82	21409	4.0250	-	7471B	PPB	30 Mar 2020 16:23:15	U	SMPL
ICB82 - 1	ICB82	-13	-0.0434	-	7471B	PPB	30 Mar 2020 16:25:30	U	SMPL
CCV24 - 1	CCV24	27864	5.2509	-	7471B	PPB	30 Mar 2020 16:27:46	U	SMPL
CCB24 - 1	CCB24	-26	-0.0459	-	7471B	PPB	30 Mar 2020 16:30:00	U	SMPL
CRA - 1	CRA	1040	0.1565	-	7471B	PPB	30 Mar 2020 16:32:16	U	SMPL
HighStd - 1	HighStd	52011	9.8369	-	7471B	PPB	30 Mar 2020 16:34:31	U	SMPL
ChkStd - 1	ChkStd	37727	7.1241	-	7471B	PPB	30 Mar 2020 16:36:46	U	SMPL
PB127884BL - 1	PBS	-216	-0.0820	-	7471B	PPB	30 Mar 2020 16:39:04	U	SMPL
PB127884BS - 1	LCSS	19925	3.7432	-	7471B	PPB	30 Mar 2020 16:41:21	U	SMPL
L1762-23 - 1	TR-04-032720	3775	0.6760	-	7471B	PPB	30 Mar 2020 16:43:37	U	SMPL
L2055-01 - 1	SB-2-(1-3)	6934	1.2759	-	7471B	PPB	30 Mar 2020 16:45:54	U	SMPL
L2055-02 - 1	SB-2-(10-12)	1120	0.1717	-	7471B	PPB	30 Mar 2020 16:48:11	U	SMPL
L2055-02DUP - 1	SB-2-(10-12)DUP	977	0.1446	-	7471B	PPB	30 Mar 2020 16:50:28	U	SMPL
L2055-02MS - 1	SB-2-(10-12)MS	21185	3.9824	-	7471B	PPB	30 Mar 2020 16:52:43	U	SMPL
CCV25 - 1	CCV25	27978	5.2726	-	7471B	PPB	30 Mar 2020 16:54:58	U	SMPL
CCB25 - 1	CCB25	-53	-0.0510	-	7471B	PPB	30 Mar 2020 16:57:15	U	SMPL
L2055-02MSD - 1	SB-2-(10-12)MSD	21714	4.0829	-	7471B	PPB	30 Mar 2020 16:59:31	U	SMPL
L2055-03 - 1	SB-6(1-3)	8801	1.6305	-	7471B	PPB	30 Mar 2020 17:01:46	U	SMPL
L2055-04 - 1	SB-6-(10-12)	981	0.1453	-	7471B	PPB	30 Mar 2020 17:04:02	U	SMPL
L2055-05 - 1	SB-8-(1-3)	5570	1.0169	-	7471B	PPB	30 Mar 2020 17:06:17	U	SMPL
L2055-06 - 1	SB-10-(1-3)	4391	0.7930	-	7471B	PPB	30 Mar 2020 17:08:33	U	SMPL
L2055-07 - 1	SB-9-(1-3)	4973	0.9035	-	7471B	PPB	30 Mar 2020 17:10:49	U	SMPL
L2055-08 - 1	SB-7-(1-3)	2655	0.4633	-	7471B	PPB	30 Mar 2020 17:13:05	U	SMPL
L2055-02LX5 - 1		163	-0.0100	-	7471B	PPB	30 Mar 2020 17:15:22	U	SMPL
L2055-02A - 1		21688	4.0780	-	7471B	PPB	30 Mar 2020 17:17:39	U	SMPL
CCV26 - 1	CCV26	28118	5.2992	-	7471B	PPB	30 Mar 2020 17:19:55	U	SMPL
CCB26 - 1	CCB26	-50	-0.0505	-	7471B	PPB	30 Mar 2020 17:22:10	U	SMPL