



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108523

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/31/2020	14:45
2	CAL2	1	Water	NA	NA	20.2	7.00	03/31/2020	14:46
3	CAL3	1	Water	NA	NA	20.2	10.03	03/31/2020	14:48
4	ICV	1	Water	NA	NA	20.1	6.98	03/31/2020	15:00
5	CCV1	1	Water	NA	NA	20.2	2.02	03/31/2020	15:01
6	L2097-10	1	Solid	20.02	20	19.4	8.72	03/31/2020	15:02
7	L2097-10DUP	1	Solid	20.03	20	19.5	8.73	03/31/2020	15:03
8	L2097-11	1	Solid	20.05	20	18.7	7.62	03/31/2020	15:05
9	L2097-17	1	Solid	20.02	20	19.8	7.73	03/31/2020	15:09
10	CCV2	1	Water	NA	NA	20.1	12.02	03/31/2020	15:10

108523

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L2074 WorkList ID : 137485 Department : Wet-Chemistry Date : 03-31-2020 14:37:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/31/2020	Solid	L2097-10 ✓	pH	Cool 4 deg C	PSEG01	G-11	40849	03/31/2020	9045D
03/31/2020	Solid	L2097-11 ✓	pH	Cool 4 deg C	PSEG01	G-11	40850	03/31/2020	9045D
03/31/2020	Solid	L2097-17 ✓	pH	Cool 4 deg C	PSEG01	G-11	40524	03/31/2020	9045D

Date/Time 03-31-2020 14:40
 Received by: Jo
 Relinquished by: Jo

Date/Time 03-31-2020 17:00
 Received by: Jo
 Relinquished by: Jo