



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108524

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/31/2020	14:45
2	CAL2	1	Water	NA	NA	20.2	7.00	03/31/2020	14:46
3	CAL3	1	Water	NA	NA	20.2	10.03	03/31/2020	14:48
4	ICV	1	Water	NA	NA	20.1	6.98	03/31/2020	15:00
5	CCV1	1	Water	NA	NA	20.2	2.01	03/31/2020	15:20
6	L2094-01	1	Water	NA	NA	17.0	5.44	03/31/2020	15:22
7	L2094-03	1	Water	NA	NA	17.1	4.65	03/31/2020	15:24
8	L2097-01	1	Water	NA	NA	14.7	6.54	03/31/2020	15:25
9	L2097-02	1	Water	NA	NA	14.7	8.95	03/31/2020	15:26
10	L2097-03	1	Water	NA	NA	15.0	8.90	03/31/2020	15:28
11	L2097-04	1	Water	NA	NA	15.1	7.71	03/31/2020	15:30
12	L2097-05	1	Water	NA	NA	16.2	7.89	03/31/2020	15:33
13	L2097-06	1	Water	NA	NA	16.9	8.47	03/31/2020	15:35
14	L2097-07	1	Water	NA	NA	16.1	8.37	03/31/2020	15:36
15	L2097-08	1	Water	NA	NA	15.9	8.99	03/31/2020	15:38
16	CCV2	1	Water	NA	NA	20.1	12.02	03/31/2020	15:42
17	L2097-09	1	Water	NA	NA	16.6	9.51	03/31/2020	15:44
18	L2097-12	1	Water	NA	NA	18.7	7.62	03/31/2020	15:45
19	L2097-13	1	Water	NA	NA	17.6	5.15	03/31/2020	15:46
20	L2097-14	1	Water	NA	NA	17.8	9.06	03/31/2020	15:48

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	L2097-15	1	Water	NA	NA	18.3	5.68	03/31/2020	15:50
22	L2097-16	1	Water	NA	NA	18.7	8.01	03/31/2020	16:00
23	L2097-16DUP	1	Water	NA	NA	18.8	8.02	03/31/2020	16:02
24	CCV3	1	Water	NA	NA	20.2	2.02	03/31/2020	16:05

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2074 WorkList ID : 137484 Department : Wet-Chemistry Date : 03-31-2020 14:36:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/31/2020	Water	L2094-01	✓ pH	Cool 4 deg C	PSEG01	G11	65-66-67	03/31/2020	9040C
03/31/2020	Water	L2094-03	✓ pH	Cool 4 deg C	PSEG01	G11	95-63-64	03/31/2020	9040C
03/31/2020	Water	L2097-01	✓ pH	Cool 4 deg C	PSEG01	G11	40731	03/31/2020	9040C
03/31/2020	Water	L2097-02	✓ pH	Cool 4 deg C	PSEG01	G11	317209	03/31/2020	9040C
03/31/2020	Water	L2097-03	✓ pH	Cool 4 deg C	PSEG01	G11	317210	03/31/2020	9040C
03/31/2020	Water	L2097-04	✓ pH	Cool 4 deg C	PSEG01	G11	317202	03/31/2020	9040C
03/31/2020	Water	L2097-05	✓ pH	Cool 4 deg C	PSEG01	G11	317203	03/31/2020	9040C
03/31/2020	Water	L2097-06	✓ pH	Cool 4 deg C	PSEG01	G11	317204	03/31/2020	9040C
03/31/2020	Water	L2097-07	✓ pH	Cool 4 deg C	PSEG01	G11	317206	03/31/2020	9040C
03/31/2020	Water	L2097-08	✓ pH	Cool 4 deg C	PSEG01	G11	317207	03/31/2020	9040C
03/31/2020	Water	L2097-09	✓ pH	Cool 4 deg C	PSEG01	G11	317208	03/31/2020	9040C
03/31/2020	Water	L2097-12	✓ pH	Cool 4 deg C	PSEG01	G11	40772	03/31/2020	9040C
03/31/2020	Water	L2097-13	✓ pH	Cool 4 deg C	PSEG01	G11	317205	03/31/2020	9040C
03/31/2020	Water	L2097-14	✓ pH	Cool 4 deg C	PSEG01	G11	40789	03/31/2020	9040C
03/31/2020	Water	L2097-15	✓ pH	Cool 4 deg C	PSEG01	G11	40788	03/31/2020	9040C
03/31/2020	Water	L2097-16	✓ pH	Cool 4 deg C	PSEG01	G11	40851	03/31/2020	9040C

Date/Time 03-31-2020 1440
 Received by: JK
 Relinquished by: JK

Date/Time 03-31-2020 17:20
 Received by: JK
 Relinquished by: JK