



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108541

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/01/2020	12:25
2	CAL2	1	Water	NA	NA	20.1	7.01	04/01/2020	12:26
3	CAL3	1	Water	NA	NA	20.1	10.03	04/01/2020	12:28
4	ICV	1	Water	NA	NA	20.2	6.99	04/01/2020	12:30
5	CCV1	1	Water	NA	NA	20.1	2.01	04/01/2020	12:32
6	L2101-04	1	Water	NA	NA	20.4	7.21	04/01/2020	12:33
7	L2101-04DUP	1	Water	NA	NA	20.5	7.22	04/01/2020	12:34
8	CCV2	1	Water	NA	NA	20.2	12.02	04/01/2020	12:35

18 (0854)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2104

WorkList ID : 137523

Department : Wet-Chemistry

Date : 04-01-2020 12:17:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2020	Water	L2101-04-Y	pH	Cool 4 deg C	PARS15	F51	239-WC-L-01	03/30/2020	9040C

Date/Time 04-01-2020 12:20
Received by: J
Relinquished by: Jh

Date/Time 04-01-2020 13:00
Received by: Jn
Relinquished by: Jn