



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108542

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/01/2020	16:00
2	CAL2	1	Water	NA	NA	20.2	7.00	04/01/2020	16:02
3	CAL3	1	Water	NA	NA	20.2	10.03	04/01/2020	16:03
4	ICV	1	Water	NA	NA	20.2	7.01	04/01/2020	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	04/01/2020	16:06
6	L2101-01	1	Solid	20.03	20	20.5	7.72	04/01/2020	16:10
7	L2101-02	1	Solid	20.04	20	20.6	8.50	04/01/2020	16:11
8	L2101-03	1	Solid	20.01	20	21.0	8.44	04/01/2020	16:13
9	L2103-06	1	Solid	20.04	20	20.4	6.51	04/01/2020	16:15
10	L2104-01	1	Solid	20.02	20	20.4	6.60	04/01/2020	16:19
11	L2104-02	1	Solid	20.03	20	20.8	6.63	04/01/2020	16:20
12	L2104-03	1	Solid	20.05	20	20.9	6.87	04/01/2020	16:22
13	L2108-06	1	Solid	20.02	20	20.8	7.56	04/01/2020	16:25
14	L2109-01	1	Solid	20.03	20	20.9	7.45	04/01/2020	16:26
15	L2109-02	1	Solid	20.05	20	20.2	6.84	04/01/2020	16:30
16	CCV2	1	Water	NA	NA	20.1	12.02	04/01/2020	16:33
17	L2122-05	1	Solid	20.02	20	22.4	8.46	04/01/2020	16:35
18	L2122-05DUP	1	Solid	20.03	20	22.5	8.47	04/01/2020	16:36
19	CCV3	1	Water	NA	NA	20.2	2.01	04/01/2020	16:38

WORKLIST(Hardcopy Internal Chain)

LB108542

WorkList Name : CORROSIONITY L2101			WorkList ID : 137522		Department : Wet-Chemistry		Date : 04-01-2020 12:16:48		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2020	Solid	L2101-01	Corrosivity	Cool 4 deg C	PARS15	F51	239-WC-S-SB-01	03/30/2020	9045D
03/30/2020	Solid	L2101-02	Corrosivity	Cool 4 deg C	PARS15	F51	239-WC-S-SB-BC	03/30/2020	9045D
03/30/2020	Solid	L2101-03	Corrosivity	Cool 4 deg C	PARS15	F51	239-WC-S-SB-020304	03/30/2020	9045D
03/30/2020	Solid	L2103-06	Corrosivity	Cool 4 deg C	PSEG01	G12	TP-60-61-COMP	03/30/2020	9045D
03/30/2020	Solid	L2104-01	Corrosivity	Cool 4 deg C	CLEA11		CHASSIE-WASH	03/30/2020	9045D
03/30/2020	Solid	L2104-02	Corrosivity	Cool 4 deg C	CLEA11		BUS-WASH	03/30/2020	9045D
03/30/2020	Solid	L2104-03	Corrosivity	Cool 4 deg C	CLEA11		OIL-WATER-SEPARATOR	03/30/2020	9045D
03/31/2020	Solid	L2108-06	Corrosivity	Cool 4 deg C	PSEG01	G12	TP-70-71-COMP	03/31/2020	9045D
03/31/2020	Solid	L2109-01	Corrosivity	Cool 4 deg C	CLEA11	H11	CHASSIE-WASH	03/31/2020	9045D
03/31/2020	Solid	L2109-02	Corrosivity	Cool 4 deg C	CLEA11	H11	OIL-WATER-SEPARATOR-1	03/31/2020	9045D
04/01/2020	Solid	L2122-05	Corrosivity	Cool 4 deg C	PSEG05	G12	SU-01-040120	04/01/2020	9045D

Date/Time 04-01-2020 15:35
 Received by: JN
 Relinquished by: DK

Date/Time 04-01-2020 17:35
 Received by: DK
 Relinquished by: JJ