



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108547

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/02/2020	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	04/02/2020	11:11
3	CAL3	1	Water	NA	NA	20.2	10.03	04/02/2020	11:12
4	ICV	1	Water	NA	NA	20.2	7.01	04/02/2020	11:15
5	CCV1	1	Water	NA	NA	20.1	2.01	04/02/2020	11:16
6	L2126-03	1	Water	NA	NA	19.9	6.40	04/02/2020	11:19
7	L2126-03DUP	1	Water	NA	NA	19.8	6.41	04/02/2020	11:20
8	L2126-04	1	Water	NA	NA	19.7	11.28	04/02/2020	11:22
9	CCV2	1	Water	NA	NA	20.1	12.02	04/02/2020	11:24

WORKLIST(Hardcopy Internal Chain)

28108547

WorkList Name: CORROSIVITY L2126 WorkList ID: 137556 Department: Wet-Chemistry Date: 04-02-2020 10:54:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/01/2020	Water	L2126-03	☒ Corrosivity	Cool 4 deg C	PSEG01	G12	2-5-1	04/01/2020	9040C
04/01/2020	Water	L2126-04	☒ Corrosivity	Cool 4 deg C	PSEG01	G12	2-4-1	04/01/2020	9040C

Date/Time 04-02-2020 11:00
 Received by: Jn
 Relinquished by: Jn

Date/Time 04-02-2020 12:10
 Received by: Jn
 Relinquished by: Jn