



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108549

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/02/2020	11:10
2	CAL2	1	Water	NA	NA	20.2	7.00	04/02/2020	11:11
3	CAL3	1	Water	NA	NA	20.2	10.03	04/02/2020	11:12
4	ICV	1	Water	NA	NA	20.2	7.01	04/02/2020	11:15
5	CCV1	1	Water	NA	NA	20.1	2.02	04/02/2020	12:00
6	L2134-01	1	Water	NA	NA	17.1	6.94	04/02/2020	12:02
7	L2134-01DUP	1	Water	NA	NA	17.2	6.95	04/02/2020	12:03
8	L2134-04	1	Water	NA	NA	17.4	6.96	04/02/2020	12:05
9	CCV2	1	Water	NA	NA	20.2	12.01	04/02/2020	12:06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2135 WorkList ID : 137558 Department : Wet-Chemistry Date : 04-02-2020 11:49:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/01/2020	Water	L2134-01	pH	Cool 4 deg C	APT01	G11	WGGM-20-WWW-TSF-012	04/01/2020	SM 4500-H B
04/01/2020	Water	L2134-04	pH	Cool 4 deg C	APT01	G11	WGGM-20-WWW-TSF-013	04/01/2020	SM 4500-H B

Date/Time 04.02.2020 11:50
 Received by: J
 Relinquished by: JH

Date/Time 04.02.2020 12:50
 Received by: J
 Relinquished by: JH