



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108566

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/03/2020	10:35
2	CAL2	1	Water	NA	NA	20.1	7.01	04/03/2020	10:36
3	CAL3	1	Water	NA	NA	20.1	10.03	04/03/2020	10:37
4	ICV	1	Water	NA	NA	20.2	6.99	04/03/2020	10:38
5	CCV1	1	Water	NA	NA	20.2	2.01	04/03/2020	10:40
6	L2143-01	1	Water	NA	NA	20.0	7.62	04/03/2020	10:42
7	L2143-01DUP	1	Water	NA	NA	20.1	7.63	04/03/2020	10:43
8	L2152-02	1	Water	NA	NA	14.6	5.64	04/03/2020	10:46
9	L2152-03	1	Water	NA	NA	18.0	8.00	04/03/2020	10:47
10	CCV2	1	Water	NA	NA	20.1	12.02	04/03/2020	10:48

WORKLIST(Hardcopy Internal Chain)

18108566

WorkList Name : PH L2143

WorkList ID : 137592

Department : Wet-Chemistry

Date : 04-03-2020 10:26:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/02/2020	Water	L2143-01	pH	Cool 4 deg C	PSEG01	G21	FORMER-OLDEN-AVE-SUB	04/02/2020	9040C
04/02/2020	Water	L2152-02	pH	Cool 4 deg C	PSEG01	G22	40734	04/02/2020	9040C
04/02/2020	Water	L2152-03	pH	Cool 4 deg C	PSEG01	G22	40656	04/02/2020	9040C

Date/Time 04.03.2020 10:30
 Received by: JN
 Relinquished by: JN

Date/Time 04.03.2020 11:35
 Received by: JN
 Relinquished by: JN