



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB108573

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP83253
COD calibration std. 100 ppm	WP83249
COD calibration std. 10 ppm	WP83247
COD calibration std. 150 ppm	WP83250
COD calibration std. 50 ppm	WP83252
COD calibration std. 0 ppm	WP83246
COD CCV std, 50ppm	WP83306
COD Digestion Vials Low Level 0-150Mg/L	W2678

Temp In (C): 148	Date In: 04/03/2020	Time In: 12:10
Temp Out (C): 149	Date Out: 04/03/2020	Time Out: 14:10

Intercept: 0.4144

Slope: 1.0094

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	04/03/2020	18:04
2	CAL2	10	1	Water	11.000	04/03/2020	18:05
3	CAL3	50	1	Water	51.000	04/03/2020	18:05
4	CAL4	100	1	Water	101.000	04/03/2020	18:06
5	CAL5	150	1	Water	152.000	04/03/2020	18:06



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB108573

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	48.133	04/03/2020	18:07
2	ICB				1	Water	0.000	-0.411	04/03/2020	18:07
3	CCV1	50			1	Water	52.000	51.105	04/03/2020	18:08
4	CCB1				1	Water	0.000	-0.411	04/03/2020	18:09
5	LB108573BL				1	Water	0.000	-0.411	04/03/2020	18:09
6	LB108573BS	50			1	Water	49.000	48.133	04/03/2020	18:10
7	L2124-01				1	Water	1.000	0.580	04/03/2020	18:10
8	L2124-01DUP				1	Water	2.000	1.571	04/03/2020	18:11
9	L2124-01MS	50			1	Water	54.000	53.087	04/03/2020	18:11
10	L2124-01MSD	50			1	Water	55.000	54.077	04/03/2020	18:12
11	L2124-02				1	Water	0.000	-0.411	04/03/2020	18:12
12	L2126-01				1	Water	6.000	5.534	04/03/2020	18:13
13	L2126-02				1	Water	12.000	11.478	04/03/2020	18:13
14	L2143-01				1	Water	52.000	51.105	04/03/2020	18:14
15	CCV2	50			1	Water	52.000	51.105	04/03/2020	18:14
16	CCB2				1	Water	0.000	-0.411	04/03/2020	18:15