



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108578

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/03/2020	14:20
2	CAL2	1	Water	NA	NA	20.1	7.01	04/03/2020	14:21
3	CAL3	1	Water	NA	NA	20.1	10.03	04/03/2020	14:22
4	ICV	1	Water	NA	NA	20.1	7.00	04/03/2020	14:25
5	CCV1	1	Water	NA	NA	20.2	2.01	04/03/2020	14:26
6	L2151-01	1	Water	NA	NA	20.7	6.99	04/03/2020	14:27
7	L2151-01DUP	1	Water	NA	NA	20.6	7.00	04/03/2020	14:29
8	L2151-02	1	Water	NA	NA	20.5	7.07	04/03/2020	14:32
9	L2151-03	1	Water	NA	NA	20.8	7.34	04/03/2020	14:35
10	L2151-04	1	Water	NA	NA	20.9	7.64	04/03/2020	14:36
11	L2151-05	1	Water	NA	NA	20.8	7.23	04/03/2020	14:38
12	L2151-06	1	Water	NA	NA	20.8	7.33	04/03/2020	14:39
13	L2151-07	1	Water	NA	NA	20.5	7.54	04/03/2020	14:40
14	L2151-08	1	Water	NA	NA	20.6	7.88	04/03/2020	14:42
15	L2151-09	1	Water	NA	NA	20.9	7.60	04/03/2020	14:45
16	CCV2	1	Water	NA	NA	20.1	12.02	04/03/2020	14:46
17	L2151-10	1	Water	NA	NA	20.5	8.26	04/03/2020	14:48
18	L2151-11	1	Water	NA	NA	20.8	7.67	04/03/2020	14:50
19	L2151-12	1	Water	NA	NA	20.8	7.61	04/03/2020	14:52
20	L2151-13	1	Water	NA	NA	20.6	7.66	04/03/2020	15:00

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	20.2	2.01	04/03/2020	15:02

WORKLIST(Hardcopy Internal Chain)

108578

WorkList Name : PH L2151

WorkList ID : 137619

Department : Wet-Chemistry

Date : 04-03-2020 14:10:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/03/2020	Water	L2151-01-1	pH	Cool 4 deg C	TULL02	G22	MH-719	04/03/2020	9040C
04/03/2020	Water	L2151-02	pH	Cool 4 deg C	TULL02	G22	MH-720	04/03/2020	9040C
04/03/2020	Water	L2151-03	pH	Cool 4 deg C	TULL02	G22	CBH-139	04/03/2020	9040C
04/03/2020	Water	L2151-04	pH	Cool 4 deg C	TULL02	G22	CBH-591	04/03/2020	9040C
04/03/2020	Water	L2151-05	pH	Cool 4 deg C	TULL02	G22	MH-726	04/03/2020	9040C
04/03/2020	Water	L2151-06	pH	Cool 4 deg C	TULL02	G22	MH-725	04/03/2020	9040C
04/03/2020	Water	L2151-07	pH	Cool 4 deg C	TULL02	G22	MH-J	04/03/2020	9040C
04/03/2020	Water	L2151-08	pH	Cool 4 deg C	TULL02	G22	MH-4	04/03/2020	9040C
04/03/2020	Water	L2151-09	pH	Cool 4 deg C	TULL02	G22	MH-109	04/03/2020	9040C
04/03/2020	Water	L2151-10	pH	Cool 4 deg C	TULL02	G22	MH-F	04/03/2020	9040C
04/03/2020	Water	L2151-11	pH	Cool 4 deg C	TULL02	G22	CBH-140	04/03/2020	9040C
04/03/2020	Water	L2151-12	pH	Cool 4 deg C	TULL02	G22	MH-5	04/03/2020	9040C
04/03/2020	Water	L2151-13	pH	Cool 4 deg C	TULL02	G22	CBH-594	04/03/2020	9040C

Date/Time 04-03-2020 14:15
 Received by: J
 Relinquished by: AM

Date/Time 04-03-2020 17:20
 Received by: AM
 Relinquished by: J