



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: Daniel

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108586

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/03/2020	09:00
2	L2073-01		1	8	100.00	>212.0	>212.0	04/03/2020	09:20
3	L2073-01DUP		1	8	100.00	>212.0	>212.0	04/03/2020	09:35
4	L2094-01		1	9	100.00	>212.0	>212.0	04/03/2020	09:55
5	L2094-03		1	8	100.00	>212.0	>212.0	04/03/2020	10:15
6	L2097-12		1	8	100.00	>212.0	>212.0	04/03/2020	10:35
7	L2097-13		1	9	100.00	>212.0	>212.0	04/03/2020	10:53
8	L2097-14		1	8	100.00	>212.0	>212.0	04/03/2020	11:15
9	L2097-15		1	8	100.00	>212.0	>212.0	04/03/2020	11:36
10	L2097-16		1	8	100.00	>212.0	>212.0	04/03/2020	11:55
11	L2097-18		1	8	100.00	>212.0	>212.0	04/03/2020	12:20
12	L2097-19		1	9	100.00	>212.0	>212.0	04/03/2020	12:45
13	L2101-04		1	8	100.00	>212.0	>212.0	04/03/2020	13:10
14	L2126-03		1	9	100.00	>212.0	>212.0	04/03/2020	13:30
15	L2126-04		1	8	100.00	>212.0	>212.0	04/03/2020	13:55
16	L2143-02		1	8	100.00	>212.0	>212.0	04/03/2020	14:22
17	L2152-02		1	8	100.00	>212.0	>212.0	04/03/2020	14:42
18	L2152-03		1	8	100.00	>212.0	>212.0	04/03/2020	15:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB108586

WorkList Name : FP040320

WorkList ID : 137617

Department : Wet-Chemistry

Date : 04-03-2020 08:11:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/06/2020	Water	L2073-01	Flash Point	Cool 4 deg C	PSEG01	F51	32272-32273-21719-COMP	03/30/2020	1010A
04/07/2020	Water	L2094-01	Flash Point	Cool 4 deg C	PSEG01	G11	65-66-67	03/31/2020	1010A
04/07/2020	Water	L2094-03	Flash Point	Cool 4 deg C	PSEG01	G11	95-63-64	03/31/2020	1010A
04/07/2020	Water	L2097-12	Flash Point	Cool 4 deg C	PSEG01	G11	40772	03/31/2020	1010A
04/07/2020	Water	L2097-13	Flash Point	Cool 4 deg C	PSEG01	G11	317205	03/31/2020	1010A
04/07/2020	Water	L2097-14	Flash Point	Cool 4 deg C	PSEG01	G11	40789	03/31/2020	1010A
04/07/2020	Water	L2097-15	Flash Point	Cool 4 deg C	PSEG01	G11	40788	03/31/2020	1010A
04/07/2020	Water	L2097-16	Flash Point	Cool 4 deg C	PSEG01	G11	40851	03/31/2020	1010A
04/07/2020	Water	L2097-18	Flash Point	Cool 4 deg C	PSEG01	G11	COMP-1	03/31/2020	1010A
04/07/2020	Water	L2097-19	Flash Point	Cool 4 deg C	PSEG01	G11	COMP-2	03/31/2020	1010A
04/13/2020	Water	L2101-04	Flash Point	Cool 4 deg C	PARS15	F51	239-WC-L-01	03/30/2020	1010A
04/06/2020	Water	L2126-03	Flash Point	Cool 4 deg C	PSEG01	G12	2-5-1	04/01/2020	1010A
04/06/2020	Water	L2126-04	Flash Point	Cool 4 deg C	PSEG01	G12	2-4-1	04/01/2020	1010A
04/16/2020	Water	L2143-02	Flash Point	Cool 4 deg C	PSEG01	G21	FORMER-OLDEN-AVE-SUB	04/02/2020	1010A
04/09/2020	Water	L2152-02	Flash Point	Cool 4 deg C	PSEG01	G22	40734	04/02/2020	1010A
04/09/2020	Water	L2152-03	Flash Point	Cool 4 deg C	PSEG01	G22	40656	04/02/2020	1010A

Date/Time 04/03/2020 8:30

Received by: DP

Relinquished by: DP

Date/Time 04/03/2020 15:20

Received by: DP

Relinquished by: DP