


LB108590

7470A

INSTRUMENT ID: LY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	1519	0.000	1519				
0.2	0.200	0.237	0.037	2213	0.0 %	2213				
2.5	2.500	2.564	0.064	10213	0.0 %	10213				
5.0	5.000	4.799	-0.201	17892	0.0 %	17892				
7.5	7.500	7.451	-0.049	27008	0.0 %	27008				
10.0	10.000	10.114	0.114	36163	0.0 %	36163				

LB108590 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Apr 2020 12:46:46

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	1519	-	0.00007470A	PPB	05 Apr 2020 13:09:50	S	Std
0.2 - 1	50.2	2213	-	0.20007470A	PPB	05 Apr 2020 13:11:57	S	Std
2.5 - 1	52.5	10213	-	2.50007470A	PPB	05 Apr 2020 13:14:05	S	Std
5.0 - 1	55	17892	-	5.00007470A	PPB	05 Apr 2020 13:16:12	S	Std
7.5 - 1	57.5	27008	-	7.50007470A	PPB	05 Apr 2020 13:23:24	S	Std
10.0 - 1	510	36163	-	10.00007470A	PPB	05 Apr 2020 13:27:59	S	Std
ICV84 - 1	ICV84	15899	4.2187	-7470A	PPB	05 Apr 2020 13:33:15	U	SMPL
ICB84 - 1	ICB84	1447	0.0140	-7470A	PPB	05 Apr 2020 13:35:20	U	SMPL
CCV30 - 1	CCV30	18440	4.9580	-7470A	PPB	05 Apr 2020 13:37:26	U	SMPL
CCB30 - 1	CCB30	1513	0.0332	-7470A	PPB	05 Apr 2020 13:39:32	U	SMPL
CRA - 1	CRA	1914	0.1498	-7470A	PPB	05 Apr 2020 13:44:04	U	SMPL
HighStd - 1	HighStd	36123	10.1028	-7470A	PPB	05 Apr 2020 13:46:10	U	SMPL
ChkStd - 1	ChkStd	28668	7.9338	-7470A	PPB	05 Apr 2020 13:48:15	U	SMPL
PB128033BL - 1	PBW	1720	0.0934	-7470A	PPB	05 Apr 2020 13:52:48	U	SMPL
PB128033BS - 1	LCWS	14626	3.8483	-7470A	PPB	05 Apr 2020 13:57:18	U	SMPL
L2117-05 - 1	OK-03-040220	742	-0.1912	-7470A	PPB	05 Apr 2020 14:13:05	U	SMPL
L2129-04 - 1	TP-9	1530	0.0381	-7470A	PPB	05 Apr 2020 14:17:35	U	SMPL
L2129-04DUP - 1	TP-9DUP	1387	-0.0035	-7470A	PPB	05 Apr 2020 14:19:42	U	SMPL
L2129-04MS - 1	TP-9MS	13295	3.4611	-7470A	PPB	05 Apr 2020 14:21:50	U	SMPL
L2129-04MSD - 1	TP-9MSD	13150	3.4189	-7470A	PPB	05 Apr 2020 14:23:56	U	SMPL
CCV31 - 1	CCV31	18221	4.8943	-7470A	PPB	05 Apr 2020 14:26:02	U	SMPL
CCB31 - 1	CCB31	1534	0.0393	-7470A	PPB	05 Apr 2020 14:28:07	U	SMPL
L2129-09 - 1	TP-8	912	-0.1417	-7470A	PPB	05 Apr 2020 14:32:36	U	SMPL
L2129-14 - 1	TP-7	733	-0.1938	-7470A	PPB	05 Apr 2020 14:34:41	U	SMPL
L2129-19 - 1	TP-4	855	-0.1583	-7470A	PPB	05 Apr 2020 14:36:47	U	SMPL
L2129-24 - 1	TP-1	774	-0.1819	-7470A	PPB	05 Apr 2020 14:41:19	U	SMPL
L2129-29 - 1	TP-2	1044	-0.1033	-7470A	PPB	05 Apr 2020 14:45:54	U	SMPL
L2129-34 - 1	TP-5	1568	0.0492	-7470A	PPB	05 Apr 2020 14:48:00	U	SMPL
L2129-39 - 1	TP-6	1165	-0.0681	-7470A	PPB	05 Apr 2020 14:50:07	U	SMPL
L2129-44 - 1	TP-3	1151	-0.0722	-7470A	PPB	05 Apr 2020 14:52:15	U	SMPL
L2132-03 - 1	T-1	1724	0.0945	-7470A	PPB	05 Apr 2020 14:54:22	U	SMPL
L2132-08 - 1	T-2	1097	-0.0879	-7470A	PPB	05 Apr 2020 14:59:10	U	SMPL
CCV32 - 1	CCV32	18439	4.9577	-7470A	PPB	05 Apr 2020 15:01:16	U	SMPL
CCB32 - 1	CCB32	1542	0.0416	-7470A	PPB	05 Apr 2020 15:03:22	U	SMPL
L2132-13 - 1	CLEAN-GRAVEL	1034	-0.1062	-7470A	PPB	05 Apr 2020 15:05:28	U	SMPL
L2142-01 - 1	TOPSOIL-WC-1	1738	0.0986	-7470A	PPB	05 Apr 2020 15:07:35	U	SMPL
L2144-01 - 1	CHASSIS-WASH	1170	-0.0666	-7470A	PPB	05 Apr 2020 15:14:47	U	SMPL
L2144-02 - 1	BUS-WASH	1259	-0.0407	-7470A	PPB	05 Apr 2020 15:19:22	U	SMPL
L2144-03 - 1	RECLAIM-SYSTEM	1533	0.0390	-7470A	PPB	05 Apr 2020 15:21:28	U	SMPL
L2146-06 - 1	TP-67-66-COMP	1728	0.0957	-7470A	PPB	05 Apr 2020 15:23:35	U	SMPL
L2146-13 - 1	TP-65-COMP	1702	0.0881	-7470A	PPB	05 Apr 2020 15:25:43	U	SMPL
PB128033TB - 1	PB127980TB	1092	-0.0893	-7470A	PPB	05 Apr 2020 15:27:50	U	SMPL
L2129-04LX5 - 1		1139	-0.0757	-7470A	PPB	05 Apr 2020 15:29:58	U	SMPL
L2129-04A - 1		13166	3.4236	-7470A	PPB	05 Apr 2020 15:32:05	U	SMPL
CCV33 - 1	CCV33	18429	4.9548	-7470A	PPB	05 Apr 2020 15:34:11	U	SMPL
CCB33 - 1	CCB33	1567	0.0489	-7470A	PPB	05 Apr 2020 15:36:17	U	SMPL