


LB108595

7471B

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	-28	0.000	-28				
0.2	0.200	0.197	-0.003	1035	0.0 %	1035				
2.5	2.500	2.420	-0.080	13715	0.0 %	13715				
5.0	5.000	5.044	0.044	28689	0.0 %	28689				
7.5	7.500	7.636	0.136	43474	0.0 %	43474				
10.0	10.000	9.892	-0.108	56345	0.0 %	56345				

LB108595 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 06 Apr 2020 11:46:31				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-28	-	0.00007471B	PPB	06 Apr 2020 11:57:32	S	Std
0.2 - 1	50.2	1035	-	0.20007471B	PPB	06 Apr 2020 11:59:49	S	Std
2.5 - 1	52.5	13715	-	2.50007471B	PPB	06 Apr 2020 12:02:06	S	Std
5.0 - 1	55	28689	-	5.00007471B	PPB	06 Apr 2020 12:06:56	S	Std
7.5 - 1	57.5	43474	-	7.50007471B	PPB	06 Apr 2020 12:09:12	S	Std
10.0 - 1	510	56345	-	10.00007471B	PPB	06 Apr 2020 12:11:29	S	Std
ICV85 - 1	ICV85	22325	3.9289	-7471B	PPB	06 Apr 2020 12:14:13	U	SMPL
ICB85 - 1	ICB85	-237	-0.0259	-7471B	PPB	06 Apr 2020 12:16:29	U	SMPL
CCV34 - 1	CCV34	27290	4.7992	-7471B	PPB	06 Apr 2020 12:18:45	U	SMPL
CCB34 - 1	CCB34	-261	-0.0301	-7471B	PPB	06 Apr 2020 12:21:00	U	SMPL
CRA - 1	CRA	1035	0.1971	-7471B	PPB	06 Apr 2020 12:23:16	U	SMPL
HighStd - 1	HighStd	56869	9.9839	-7471B	PPB	06 Apr 2020 12:25:31	U	SMPL
ChkStd - 1	ChkStd	38544	6.7718	-7471B	PPB	06 Apr 2020 12:27:46	U	SMPL
PB128038BL - 1	PBS	-354	-0.0464	-7471B	PPB	06 Apr 2020 12:30:04	U	SMPL
PB128038BS - 1	LCSS	22663	3.9881	-7471B	PPB	06 Apr 2020 12:32:21	U	SMPL
L2102-01 - 1	RT-46-MEDIAN	1276	0.2393	-7471B	PPB	06 Apr 2020 12:34:37	U	SMPL
L2108-05 - 1	TP-70-71-COMP	3235	0.5827	-7471B	PPB	06 Apr 2020 12:36:55	U	SMPL
L2110-01 - 1	AU-05-040220	3898	0.6989	-7471B	PPB	06 Apr 2020 12:39:11	U	SMPL
L2110-01DUP - 1	AU-05-040220DUP	3686	0.6618	-7471B	PPB	06 Apr 2020 12:41:28	U	SMPL
L2110-01MS - 1	AU-05-040220MS	24193	4.2563	-7471B	PPB	06 Apr 2020 12:43:43	U	SMPL
CCV35 - 1	CCV35	27288	4.7988	-7471B	PPB	06 Apr 2020 12:45:58	U	SMPL
CCB35 - 1	CCB35	-167	-0.0136	-7471B	PPB	06 Apr 2020 12:48:14	U	SMPL
L2110-01MSD - 1	AU-05-040220MSD	24442	4.3000	-7471B	PPB	06 Apr 2020 12:50:30	U	SMPL
L2112-01 - 1	HR-06-040220	767	0.1501	-7471B	PPB	06 Apr 2020 12:52:46	U	SMPL
L2117-01 - 1	OK-03-040220	860	0.1664	-7471B	PPB	06 Apr 2020 12:55:02	U	SMPL
L2122-01 - 1	SU-01-040120	2172	0.3964	-7471B	PPB	06 Apr 2020 12:57:17	U	SMPL
L2146-05 - 1	TP-67-66-COMP	1068	0.2029	-7471B	PPB	06 Apr 2020 12:59:33	U	SMPL
L2146-12 - 1	TP-65-COMP	1877	0.3447	-7471B	PPB	06 Apr 2020 13:01:49	U	SMPL
L2153-02 - 1	VNJ-231	1145	0.2164	-7471B	PPB	06 Apr 2020 13:04:05	U	SMPL
L2153-04 - 1	72-11991	1359	0.2539	-7471B	PPB	06 Apr 2020 13:06:22	U	SMPL
L2153-06 - 1	RBR-2520699	4978	0.8882	-7471B	PPB	06 Apr 2020 13:08:39	U	SMPL
L2154-01 - 1	VNJ-212	2468	0.4483	-7471B	PPB	06 Apr 2020 13:10:55	U	SMPL
CCV36 - 1	CCV36	27727	4.8758	-7471B	PPB	06 Apr 2020 13:13:10	U	SMPL
CCB36 - 1	CCB36	-160	-0.0124	-7471B	PPB	06 Apr 2020 13:15:25	U	SMPL
L2157-01 - 1	VNJ-226	1170	0.2207	-7471B	PPB	06 Apr 2020 13:17:41	U	SMPL
L2162-05 - 1	TP-54-53-COMP	853	0.1652	-7471B	PPB	06 Apr 2020 13:19:57	U	SMPL
L2162-12 - 1	TP-50-COMP	2363	0.4299	-7471B	PPB	06 Apr 2020 13:22:12	U	SMPL
L2164-02 - 1	ROLL-OFF-COMP	1130	0.2137	-7471B	PPB	06 Apr 2020 13:24:27	U	SMPL
L2166-05 - 1	TP-47-COMP	2736	0.4952	-7471B	PPB	06 Apr 2020 13:26:43	U	SMPL
L2166-12 - 1	TP-52-46-COMP	697	0.1378	-7471B	PPB	06 Apr 2020 13:28:59	U	SMPL
L2166-19 - 1	TP-45-COMP	840	0.1629	-7471B	PPB	06 Apr 2020 13:31:15	U	SMPL
PB128055BL - 1	PBS	-55	0.0060	-7471B	PPB	06 Apr 2020 13:33:32	U	SMPL
PB128055BS - 1	LCSS	22379	3.9384	-7471B	PPB	06 Apr 2020 13:35:49	U	SMPL
L2129-01 - 1	TP-9	141924	24.8928	-7471B	PPB	06 Apr 2020 13:38:05	U	SMPL
CCV37 - 1	CCV37	27411	4.8204	-7471B	PPB	06 Apr 2020 13:40:20	U	SMPL
CCB37 - 1	CCB37	-204	-0.0201	-7471B	PPB	06 Apr 2020 13:42:40	U	SMPL
L2129-01DUP - 1	TP-9DUP	148217	25.9958	-7471B	PPB	06 Apr 2020 13:44:57	U	SMPL
L2129-01MS - 1	TP-9MS	148504	26.0461	-7471B	PPB	06 Apr 2020 13:47:12	U	SMPL
L2129-01MSD - 1	TP-9MSD	149590	26.2365	-7471B	PPB	06 Apr 2020 13:49:31	U	SMPL
L2129-06 - 1	TP-8	66359	11.6474	-7471B	PPB	06 Apr 2020 13:51:52	U	SMPL
L2129-11 - 1	TP-7	217399	38.1224	-7471B	PPB	06 Apr 2020 13:54:11	U	SMPL
L2129-16 - 1	TP-4	33577	5.9012	-7471B	PPB	06 Apr 2020 13:56:30	U	SMPL
L2129-21 - 1	TP-1	17004	2.9962	-7471B	PPB	06 Apr 2020 13:58:58	U	SMPL
L2129-26 - 1	TP-2	15596	2.7494	-7471B	PPB	06 Apr 2020 14:01:16	U	SMPL
L2129-31 - 1	TP-5	122516	21.4908	-7471B	PPB	06 Apr 2020 14:03:34	U	SMPL
L2129-36 - 1	TP-6	79910	14.0227	-7471B	PPB	06 Apr 2020 14:05:50	U	SMPL
CCV38 - 1	CCV38	27614	4.8560	-7471B	PPB	06 Apr 2020 14:08:10	U	SMPL
CCB38 - 1	CCB38	-214	-0.0218	-7471B	PPB	06 Apr 2020 14:10:27	U	SMPL
L2129-41 - 1	TP-3	38840	6.8237	-7471B	PPB	06 Apr 2020 14:12:44	U	SMPL
L2132-01 - 1	T-1	-274	-0.0324	-7471B	PPB	06 Apr 2020 14:15:00	U	SMPL
L2132-06 - 1	T-2	244	0.0584	-7471B	PPB	06 Apr 2020 14:17:17	U	SMPL
L2132-11 - 1	CLEAN-GRAVEL	-59	0.0053	-7471B	PPB	06 Apr 2020 14:19:32	U	SMPL
L2133-01 - 1	BIG-HOUSE-FLOORS	917	0.1764	-7471B	PPB	06 Apr 2020 14:21:48	U	SMPL
L2133-02 - 1	BIG-HOUSE-WALLS	5194	0.9261	-7471B	PPB	06 Apr 2020 14:24:04	U	SMPL
L2133-03 - 1	SMALL-HOUSE-FLOORS	-57	0.0057	-7471B	PPB	06 Apr 2020 14:26:21	U	SMPL
L2133-04 - 1	SMALL-HOUSE-WALLS	5668	1.0092	-7471B	PPB	06 Apr 2020 14:28:39	U	SMPL
L2147-01 - 1	FK-GF-02-056-A	1543	0.2861	-7471B	PPB	06 Apr 2020 14:30:56	U	SMPL
L2147-03 - 1	FK-GF-02-056-B	305	0.0691	-7471B	PPB	06 Apr 2020 14:33:13	U	SMPL
CCV39 - 1	CCV39	25797	4.5375	-7471B	PPB	06 Apr 2020 14:47:59	U	SMPL
CCB39 - 1	CCB39	-155	-0.0115	-7471B	PPB	06 Apr 2020 14:50:14	U	SMPL
L2148-01 - 1	FK-GF-02-055-A	8081	1.4321	-7471B	PPB	06 Apr 2020 14:52:30	U	SMPL
L2148-03 - 1	FK-GF-02-055-B	5000	0.8921	-7471B	PPB	06 Apr 2020 14:54:47	U	SMPL
L2110-01LX5 - 1		622	0.1247	-7471B	PPB	06 Apr 2020 14:57:03	U	SMPL
L2110-01A - 1		23614	4.1548	-7471B	PPB	06 Apr 2020 14:59:19	U	SMPL
L2129-01LX5 - 1		27415	4.8211	-7471B	PPB	06 Apr 2020 15:01:35	U	SMPL
L2129-01A - 1		149534	26.2267	-7471B	PPB	06 Apr 2020 15:03:52	U	SMPL
L2129-06DLX2 - 1		31568	5.5490	-7471B	PPB	06 Apr 2020 15:06:09	U	SMPL
L2129-11DLX10 - 1		21676	3.8151	-7471B	PPB	06 Apr 2020 15:08:29	U	SMPL
L2129-31DLX5 - 1		23695	4.1690	-7471B	PPB	06 Apr 2020 15:10:47	U	SMPL
L2129-36DLX5 - 1		15339	2.7044	-7471B	PPB	06 Apr 2020 15:13:05	U	SMPL
CCV40 - 1	CCV40	26499	4.6605	-7471B	PPB	06 Apr 2020 15:15:22	U	SMPL

LB108595 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Apr 2020 11:46:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB40 - 1	CCB40	-181	-0.0161		-7471B	PPB	06 Apr 2020 15:17:38	U	SMPL
L2129-01DLC5 - 1		26174	4.6036		-7471B	PPB	06 Apr 2020 15:19:55	U	SMPL
L2129-01DUPDLX5 - 1		26872	4.7259		-7471B	PPB	06 Apr 2020 15:22:11	U	SMPL
L2129-01MSDLX5 - 1		26934	4.7368		-7471B	PPB	06 Apr 2020 15:24:29	U	SMPL
L2129-01MSDDLX5 - 1		25620	4.5065		-7471B	PPB	06 Apr 2020 15:26:46	U	SMPL
L2129-01LDLX25 - 1		5263	0.9382		-7471B	PPB	06 Apr 2020 15:29:03	U	SMPL
L2129-01ADLX5 - 1		32266	5.6714		-7471B	PPB	06 Apr 2020 15:31:21	U	SMPL
CCV41 - 1	CCV41	26840	4.7203		-7471B	PPB	06 Apr 2020 15:36:14	U	SMPL
CCB41 - 1	CCB41	-188	-0.0173		-7471B	PPB	06 Apr 2020 15:38:30	U	SMPL