



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 4/7/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 14:50  
In Date: 04/06/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 07:32  
Out Date: 04/07/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB108596

| Lab ID   | Client SampleID | Dish # | Dish Wt(g) (A) | Dish + Sample Wt(g) (B) | Dish+Dry Sample Wt(g) (C) | % Solid | Comments    |
|----------|-----------------|--------|----------------|-------------------------|---------------------------|---------|-------------|
| L2155-01 | DBF98           | 1      | 1.00           | 2.00                    | 2.00                      | 100.0   | P.E. SAMPLE |
| L2155-02 | DBFA1           | 2      | 1.15           | 9.76                    | 8.29                      | 82.9    |             |
| L2155-03 | DBFA2           | 3      | 1.15           | 9.96                    | 8.54                      | 83.9    |             |
| L2155-04 | DBFA3           | 4      | 1.18           | 9.69                    | 7.95                      | 79.6    |             |
| L2155-05 | DBFA3MS         | 5      | 1.18           | 9.69                    | 7.95                      | 79.6    |             |
| L2155-06 | DBFA3MSD        | 6      | 1.18           | 9.69                    | 7.95                      | 79.6    |             |
| L2155-07 | DBFA4           | 7      | 1.15           | 9.96                    | 8.66                      | 85.2    |             |
| L2155-08 | DBFA5           | 8      | 1.14           | 9.86                    | 8.18                      | 80.7    |             |
| L2155-09 | DBFA6           | 9      | 1.12           | 9.92                    | 8.01                      | 78.3    |             |
| L2155-10 | DBFA7           | 10     | 1.16           | 9.97                    | 8.26                      | 80.6    |             |
| L2155-11 | DBFA8           | 11     | 1.18           | 9.58                    | 8.19                      | 83.5    |             |
| L2155-12 | DBFA9           | 12     | 1.18           | 9.68                    | 8.32                      | 84.0    |             |
| L2155-13 | DBFB0           | 13     | 1.14           | 9.65                    | 8.27                      | 83.8    |             |
| L2155-14 | DBFB1           | 14     | 1.13           | 9.87                    | 8.18                      | 80.7    |             |
| L2155-15 | DBFB2           | 15     | 1.14           | 9.9                     | 8.39                      | 82.8    |             |
| L2155-16 | DBFB3           | 16     | 1.14           | 9.77                    | 7.19                      | 70.1    |             |
| L2155-17 | DBFB4           | 17     | 1.19           | 9.93                    | 8.17                      | 79.9    |             |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$