



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108610

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/06/2020	14:20
2	CAL2	1	Water	NA	NA	20.1	7.01	04/06/2020	14:21
3	CAL3	1	Water	NA	NA	20.1	10.03	04/06/2020	14:22
4	ICV	1	Water	NA	NA	20.2	6.99	04/06/2020	14:25
5	CCV1	1	Water	NA	NA	20.1	2.01	04/06/2020	14:26
6	L2171-02	1	Water	NA	NA	19.5	7.31	04/06/2020	14:30
7	L2171-02DUP	1	Water	NA	NA	19.6	7.32	04/06/2020	14:31
8	CCV2	1	Water	NA	NA	20.2	12.02	04/06/2020	14:33

✓ 108610

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2171 WorkList ID : 137665 Department : Wet-Chemistry Date : 04-06-2020 14:06:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/06/2020	Water	L2171-02	-A Corrosivity	Cool 4 deg C	PSEG01	H11	FRAC-COMP	04/06/2020	9040C

Date/Time 04.06.2020 14:10
 Received by: CP
 Relinquished by: _____

Date/Time 04.06.2020 15:15
 Received by: CP
 Relinquished by: JN