


LB108613

ISM02.4

INSTRUMENT ID: CY1



Linear

A= 0.0000e+000

B= 2.3204e-004

C= -3.8157e-002

Rho= 0.9999762

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	-0.027	-0.027	46	0.000	46					
0.05	0.050					149					
0.20	0.200	0.206	0.006	1052	0.0 %	1052					
2.50	2.500	2.497	-0.003	10925	0.0 %	10925					
5.00	5.000	5.038	0.038	21877	0.0 %	21877					
7.50	7.500	7.521	0.021	32576	0.0 %	32576					
10.0	10.000	9.965	-0.035	43111	0.0 %	43111					

1131000

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	46	-		0 ISM02.4	PPB	4/7/2020 11:40	Std
	0.2 S0.2	1052	-		0.2 ISM02.4	PPB	4/7/2020 11:43	Std
	2.5 S2.5	10925	-		2.5 ISM02.4	PPB	4/7/2020 11:45	Std
	5 S5.0	21877	-		5 ISM02.4	PPB	4/7/2020 11:47	Std
	7.5 S7.5	32576	-		7.5 ISM02.4	PPB	4/7/2020 11:50	Std
	10 S10.0	43111	-		10 ISM02.4	PPB	4/7/2020 11:52	Std
ICV	ICV	16326	3.7502	-	ISM02.4	PPB	4/7/2020 11:55	SMPL
ICB	ICB	-58	-0.0516	-	ISM02.4	PPB	4/7/2020 11:57	SMPL
CCV048	CCV048	22202	5.1137	-	ISM02.4	PPB	4/7/2020 11:59	SMPL
CCB048	CCB048	-59	-0.0518	-	ISM02.4	PPB	4/7/2020 12:01	SMPL
L2091-01	MCOAR9	246	0.0189	-	ISM02.4	PPB	4/7/2020 12:06	SMPL
L2091-02	MCOAS8	197	0.0076	-	ISM02.4	PPB	4/7/2020 12:08	SMPL
L2091-03	MCOAT0	208	0.0101	-	ISM02.4	PPB	4/7/2020 12:10	SMPL
L2091-04	MCOAT2	218	0.0124	-	ISM02.4	PPB	4/7/2020 12:13	SMPL
L2091-05	MCOAT4	177	0.0029	-	ISM02.4	PPB	4/7/2020 12:15	SMPL
L2091-06	MCOAT6	164	-0.0001	-	ISM02.4	PPB	4/7/2020 12:17	SMPL
L2091-07	MCOAT8	180	0.0036	-	ISM02.4	PPB	4/7/2020 12:19	SMPL
L2091-08	MCOAT8D	177	0.0029	-	ISM02.4	PPB	4/7/2020 12:22	SMPL
L2091-09	MCOAT8S	4591	1.0272	-	ISM02.4	PPB	4/7/2020 12:24	SMPL
L2091-10	MCOAW0	147	-0.004	-	ISM02.4	PPB	4/7/2020 12:26	SMPL
L2091-11	MCOAW2	146	-0.0043	-	ISM02.4	PPB	4/7/2020 12:28	SMPL
PB127949BL	PBW949	231	0.0154	-	ISM02.4	PPB	4/7/2020 12:31	SMPL
PB127950BL	PBW950	107	-0.0133	-	ISM02.4	PPB	4/7/2020 12:34	SMPL
L2092-01	MCOAS0	159	-0.0013	-	ISM02.4	PPB	4/7/2020 12:36	SMPL
L2092-02	MCOAS9	142	-0.0052	-	ISM02.4	PPB	4/7/2020 12:39	SMPL
L2092-03	MCOAT1	118	-0.0108	-	ISM02.4	PPB	4/7/2020 12:41	SMPL
L2092-04	MCOAT3	120	-0.0103	-	ISM02.4	PPB	4/7/2020 12:43	SMPL
L2092-05	MCOAT5	122	-0.0098	-	ISM02.4	PPB	4/7/2020 12:45	SMPL
L2092-06	MCOAT7	137	-0.0064	-	ISM02.4	PPB	4/7/2020 12:48	SMPL
L2092-07	MCOAT9	121	-0.0101	-	ISM02.4	PPB	4/7/2020 12:50	SMPL
CCV049	CCV049	21901	5.0438	-	ISM02.4	PPB	4/7/2020 12:52	SMPL
CCB049	CCB049	-57	-0.0514	-	ISM02.4	PPB	4/7/2020 12:54	SMPL
L2092-08	MCOAT9D	90	-0.0173	-	ISM02.4	PPB	4/7/2020 12:57	SMPL
L2092-09	MCOAT9S	4483	1.0021	-	ISM02.4	PPB	4/7/2020 12:59	SMPL
L2092-10	MCOAW1	90	-0.0173	-	ISM02.4	PPB	4/7/2020 13:01	SMPL
L2092-11	MCOAW3	116	-0.0112	-	ISM02.4	PPB	4/7/2020 13:03	SMPL
PB127951BL	PBW951	64	-0.0233	-	ISM02.4	PPB	4/7/2020 13:06	SMPL
L2047-01	MBF7J0	31	-0.031	-	ISM02.4	PPB	4/7/2020 13:08	SMPL
L2047-02	MBF7J1	50	-0.0266	-	ISM02.4	PPB	4/7/2020 13:10	SMPL
L2047-03	MBF7J2	46	-0.0275	-	ISM02.4	PPB	4/7/2020 13:12	SMPL
L2047-04	MBF7J2D	23	-0.0328	-	ISM02.4	PPB	4/7/2020 13:15	SMPL
L2047-05	MBF7J2S	4744	1.0627	-	ISM02.4	PPB	4/7/2020 13:17	SMPL
PB128036BL	PBW036	2	-0.0377	-	ISM02.4	PPB	4/7/2020 13:19	SMPL
L2159-01	MCOEY9	-95	-0.0602	-	ISM02.4	PPB	4/7/2020 13:22	SMPL
L2159-02	MCOEZ0	109	-0.0129	-	ISM02.4	PPB	4/7/2020 13:24	SMPL
L2159-03	MCOEZ1	783	0.1435	-	ISM02.4	PPB	4/7/2020 13:26	SMPL

LB108613

INSTRUMENT ID : CV1

L2159-04	MCOEZ5	99	-0.0152 -	ISM02.4	PPB	4/7/2020 13:28 SMPL
L2159-05	MCOF02	-41	-0.0477 -	ISM02.4	PPB	4/7/2020 13:31 SMPL
L2159-06	MCOF03	210	0.0106 -	ISM02.4	PPB	4/7/2020 13:33 SMPL
L2159-07	MCOF03D	209	0.0103 -	ISM02.4	PPB	4/7/2020 13:35 SMPL
L2159-08	MCOF03S	4653	1.0415 -	ISM02.4	PPB	4/7/2020 13:37 SMPL
L2159-09	MCOF04	-20	-0.0428 -	ISM02.4	PPB	4/7/2020 13:40 SMPL
CCV050	CCV050	21157	4.8712 -	ISM02.4	PPB	4/7/2020 13:42 SMPL
CCB050	CCB050	-29	-0.0449 -	ISM02.4	PPB	4/7/2020 13:44 SMPL
L2159-10	MCOF05	-17	-0.0421 -	ISM02.4	PPB	4/7/2020 13:47 SMPL
L2159-11	MCOF06	-133	-0.069 -	ISM02.4	PPB	4/7/2020 13:49 SMPL
CCV051	CCV051	21280	4.8997 -	ISM02.4	PPB	4/7/2020 13:53 SMPL
CCB051	CCB051	32	-0.0307 -	ISM02.4	PPB	4/7/2020 13:55 SMPL