



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108628

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/07/2020	15:35
2	CAL2	1	Water	NA	NA	20.1	7.01	04/07/2020	15:36
3	CAL3	1	Water	NA	NA	20.1	10.03	04/07/2020	15:37
4	ICV	1	Water	NA	NA	20.1	7.01	04/07/2020	15:39
5	CCV1	1	Water	NA	NA	20.2	2.01	04/07/2020	15:40
6	L2185-01	1	Water	NA	NA	20.9	7.73	04/07/2020	15:42
7	L2185-17	1	Water	NA	NA	20.6	6.24	04/07/2020	15:43
8	L2185-18	1	Water	NA	NA	20.9	6.36	04/07/2020	15:45
9	L2186-12	1	Water	NA	NA	21.4	7.14	04/07/2020	15:46
10	L2186-13	1	Water	NA	NA	21.3	7.07	04/07/2020	15:49
11	L2186-14	1	Water	NA	NA	21.2	4.93	04/07/2020	15:50
12	L2186-15	1	Water	NA	NA	21.1	4.90	04/07/2020	15:53
13	L2186-16	1	Water	NA	NA	21.2	7.14	04/07/2020	16:00
14	L2186-17	1	Water	NA	NA	21.1	8.77	04/07/2020	16:02
15	L2186-18	1	Water	NA	NA	21.2	8.82	04/07/2020	16:05
16	CCV2	1	Water	NA	NA	20.2	12.02	04/07/2020	16:06
17	L2186-18DUP	1	Water	NA	NA	21.3	8.83	04/07/2020	16:08
18	CCV3	1	Water	NA	NA	20.1	2.02	04/07/2020	16:10

WORKLIST(Hardcopy Internal Chain)

73108628

WorkList Name : PH L2186 W WorkList ID : 137698 Department : Wet-Chemistry Date : 04-07-2020 15:20:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/07/2020	Water	L2185-01	pH	Cool 4 deg C	PSEG01	H12	30773	04/07/2020	9040C
04/07/2020	Water	L2185-17	pH	Cool 4 deg C	PSEG01	H12	30928-PH-1	04/07/2020	9040C
04/07/2020	Water	L2185-18	pH	Cool 4 deg C	PSEG01	H12	DRIP-LIQUID-PH-2	04/07/2020	9040C
04/07/2020	Water	L2186-12	pH	Cool 4 deg C	PSEG01	H12	PH-1-31019	04/07/2020	9040C
04/07/2020	Water	L2186-13	pH	Cool 4 deg C	PSEG01	H12	PH-2-31021	04/07/2020	9040C
04/07/2020	Water	L2186-14	pH	Cool 4 deg C	PSEG01	H12	PH-3-30922	04/07/2020	9040C
04/07/2020	Water	L2186-15	pH	Cool 4 deg C	PSEG01	H12	PH-4-31018	04/07/2020	9040C
04/07/2020	Water	L2186-16	pH	Cool 4 deg C	PSEG01	H12	PH-5-31017	04/07/2020	9040C
04/07/2020	Water	L2186-17	pH	Cool 4 deg C	PSEG01	H12	PH-6-30895	04/07/2020	9040C
04/07/2020	Water	L2186-18	pH	Cool 4 deg C	PSEG01	H12	PH-7-30939	04/07/2020	9040C

Date/Time 04-07-2020 15:25
 Received by: Jn
 Relinquished by: Dr

Date/Time 04-07-2020 18:20
 Received by: Dr
 Relinquished by: Jo