


LB108636

ISM02.4

INSTRUMENT ID: CV4



A= 0.0000e+000

B= 2.3088e-004

C= -1.3515e-002

Rho= 0.9998683

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.011	-0.011	10	0.000	10					14
0.05	0.050					149					0
0.20	0.200	0.228	0.028	1046	0.0 %	1046					-2
2.50	2.500	2.511	0.011	10936	0.0 %	10936					1
5.00	5.000	4.904	-0.096	21300	0.0 %	21300					0
7.50	7.500	7.601	0.101	32980	0.0 %	32980					
10.0	10.000	9.967	-0.033	43226	0.0 %	43226					

LB108636
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	10	-		0 ISM02.4	PPB	4/8/2020 16:46	Std
	0.2 S0.2	1046	-		0.2 ISM02.4	PPB	4/8/2020 16:49	Std
	2.5 S2.5	10936	-		2.5 ISM02.4	PPB	4/8/2020 16:51	Std
	5 S5.0	21300	-		5 ISM02.4	PPB	4/8/2020 16:53	Std
	7.5 S7.5	32980	-		7.5 ISM02.4	PPB	4/8/2020 16:56	Std
	10 S10.0	43226	-		10 ISM02.4	PPB	4/8/2020 16:58	Std
ICV	ICV	16186	3.7235	-	ISM02.4	PPB	4/8/2020 17:01	SMPL
ICB	ICB	-155	-0.0493	-	ISM02.4	PPB	4/8/2020 17:03	SMPL
CCV061	CCV061	21325	4.91	-	ISM02.4	PPB	4/8/2020 17:05	SMPL
CCB061	CCB061	-154	-0.0491	-	ISM02.4	PPB	4/8/2020 17:07	SMPL
PB128101BL	PBW101	108	0.0114	-	ISM02.4	PPB	4/8/2020 17:10	SMPL
L2105-01	MBF7J9	902	0.1947	-	ISM02.4	PPB	4/8/2020 17:12	SMPL
L2105-02	MBF7K2	-110	-0.0389	-	ISM02.4	PPB	4/8/2020 17:14	SMPL
L2105-03	MBF7K3	-76	-0.0311	-	ISM02.4	PPB	4/8/2020 17:17	SMPL
L2105-04	MBF7K4	-93	-0.035	-	ISM02.4	PPB	4/8/2020 17:19	SMPL
L2105-05	MBF7K5	135	0.0177	-	ISM02.4	PPB	4/8/2020 17:21	SMPL
L2105-06	MBF7K6	38	-0.0047	-	ISM02.4	PPB	4/8/2020 17:23	SMPL
L2105-07	MBF7K7	109	0.0117	-	ISM02.4	PPB	4/8/2020 17:26	SMPL
L2105-08	MBF7K8	62	0.0008	-	ISM02.4	PPB	4/8/2020 17:28	SMPL
L2105-09	MBF7M3	23	-0.0082	-	ISM02.4	PPB	4/8/2020 17:30	SMPL
L2105-10	MBF7M4	105	0.0107	-	ISM02.4	PPB	4/8/2020 17:32	SMPL
L2105-11	MBG2S8	103	0.0103	-	ISM02.4	PPB	4/8/2020 17:35	SMPL
L2105-12	MBG2S9	73	0.0033	-	ISM02.4	PPB	4/8/2020 17:37	SMPL
L2105-13	MBG2T0	110	0.0119	-	ISM02.4	PPB	4/8/2020 17:39	SMPL
L2105-14	MBG2T1	49	-0.0022	-	ISM02.4	PPB	4/8/2020 17:41	SMPL
L2105-15	MBG2W1	77	0.0043	-	ISM02.4	PPB	4/8/2020 17:44	SMPL
L2105-16	MBG2W2	59	0.0001	-	ISM02.4	PPB	4/8/2020 17:46	SMPL
L2105-17	MBG2W2D	97	0.0089	-	ISM02.4	PPB	4/8/2020 17:48	SMPL
L2105-18	MBG2W2S	4316	0.983	-	ISM02.4	PPB	4/8/2020 17:50	SMPL
L2105-19	MBG2W3	50	-0.002	-	ISM02.4	PPB	4/8/2020 17:53	SMPL
CCV062	CCV062	21193	4.8796	-	ISM02.4	PPB	4/8/2020 17:55	SMPL
CCB062	CCB062	-145	-0.047	-	ISM02.4	PPB	4/8/2020 17:57	SMPL
L2105-20	MBG2W4	18	-0.0094	-	ISM02.4	PPB	4/8/2020 18:00	SMPL
PB128102BL	PBW102	144	0.0197	-	ISM02.4	PPB	4/8/2020 18:02	SMPL
L2174-01	MBFS68	158	0.023	-	ISM02.4	PPB	4/8/2020 18:04	SMPL
L2174-02	MBFS68D	142	0.0193	-	ISM02.4	PPB	4/8/2020 18:06	SMPL
L2174-03	MBFS68S	4441	1.0118	-	ISM02.4	PPB	4/8/2020 18:09	SMPL
L2174-04	MBFS69	-115	-0.0401	-	ISM02.4	PPB	4/8/2020 18:11	SMPL
L2174-05	MBFS70	170	0.0257	-	ISM02.4	PPB	4/8/2020 18:13	SMPL
PB128103BL	PBW103	208	0.0345	-	ISM02.4	PPB	4/8/2020 18:15	SMPL
L2187-01	MBFS72	-79	-0.0318	-	ISM02.4	PPB	4/8/2020 18:18	SMPL
L2187-02	MBFS75	-102	-0.0371	-	ISM02.4	PPB	4/8/2020 18:20	SMPL
L2187-03	MBFS75D	-72	-0.0301	-	ISM02.4	PPB	4/8/2020 18:22	SMPL
L2187-04	MBFS75S	4346	0.9899	-	ISM02.4	PPB	4/8/2020 18:24	SMPL
CCV063	CCV063	20774	4.7828	-	ISM02.4	PPB	4/8/2020 18:27	SMPL
CCB063	CCB063	-184	-0.056	-	ISM02.4	PPB	4/8/2020 18:29	SMPL