



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB108640

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP83375
Chlorine Calibration std, 0.1ppm	WP83370
Chlorine Calibration std, 0.2ppm	WP83371
Chlorine Calibration std, 0.8ppm	WP83372
Chlorine Calibration std, 0.0ppm	WP83369
Chlorine Calibration std, 1.6ppm	WP83373
Residual Chlorine Calibration and CCV std, 0	WP83374

Intercept: 0.0111

Slope: 1.0655

Regression: 0.999906

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	04/09/2020	12:10
2	CAL2	0.1	1	0.000	0.120	0.120	04/09/2020	12:11
3	CAL3	0.2	1	0.000	0.220	0.220	04/09/2020	12:12
4	CAL4	0.4	1	0.000	0.450	0.450	04/09/2020	12:14
5	CAL5	0.8	1	0.000	0.870	0.870	04/09/2020	12:15
6	CAL6	1.6	1	0.000	1.710	1.710	04/09/2020	12:17



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB108640

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4300	0.4300	0.3930	04/09/2020	12:18
2	ICB				1	0.0000	0.0000	0.0000	-0.0100	04/09/2020	12:19
3	CCV1			0.4	1	0.0000	0.4500	0.4500	0.4120	04/09/2020	12:20
4	CCB1				1	0.0000	0.0000	0.0000	-0.0100	04/09/2020	12:22
5	LB108640BL	50	50		1	0.0000	0.0000	0.0000	-0.0100	04/09/2020	12:23
6	LB108640BS	50	50	0.4	1	0.0000	0.4300	0.4300	0.3930	04/09/2020	12:25
7	L2098-01	50	50		1	0.0200	0.0400	0.0200	0.0080	04/09/2020	12:26
8	L2098-01DUP	50	50		1	0.0200	0.0500	0.0300	0.0180	04/09/2020	12:27
9	CCV2			0.4	1	0.0000	0.4400	0.4400	0.4030	04/09/2020	12:29
10	CCB2				1	0.0000	0.0000	0.0000	-0.0100	04/09/2020	12:30

LB108640

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ResChlorine-4092020

WorkList ID : 137749

Department : Wet-Chemistry

Date : 04-09-2020 09:45:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2020	Water	L2098-01	Residual Chlorine	Cool 4 deg C	STER17	F61	JS9600	03/30/2020	SM4500 Cl G

Date/Time 4/09/2020 10:45
Received by: Am
Relinquished by: CP

Date/Time 4/09/2020 17:55
Received by: CP
Relinquished by: Am