



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108641

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
PH 4 BUFFER SOLUTION	W2629
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.1	4.01	04/09/2020	11:02
2	CAL2	1	Water	NA	NA	22.1	7.01	04/09/2020	11:03
3	CAL3	1	Water	NA	NA	22.1	10.04	04/09/2020	11:04
4	ICV	1	Water	NA	NA	22.1	6.98	04/09/2020	11:07
5	CCV1	1	Water	NA	NA	22.2	2.01	04/09/2020	11:09
6	L2098-01	1	Water	NA	NA	23.0	5.41	04/09/2020	11:12
7	L2098-01DUP	1	Water	NA	NA	23.0	5.39	04/09/2020	11:15
8	CCV2	1	Water	NA	NA	22.2	12.02	04/09/2020	11:17

LB108641
WORKLIST(Hardcopy Internal Chain)

WorkList Name : Ph-4092020 WorkList ID : 137746 Department : Wet-Chemistry Date : 04-09-2020 09:44:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
03/30/2020	Water	L2098-01	pH	Cool 4 deg C	STER17	F61	JS9600	03/30/2020	SM 4500-H B

Date/Time 4/09/2020 10:45
 Received by: AM
 Relinquished by: AM

Date/Time 4/09/2020 12:55
 Received by: AM
 Relinquished by: AM