



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Alexander

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108649

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.0	4.01	04/09/2020	18:07
2	CAL2	1	Water	NA	NA	22.0	7.01	04/09/2020	18:08
3	CAL3	1	Water	NA	NA	22.0	10.04	04/09/2020	18:09
4	ICV	1	Water	NA	NA	22.1	6.99	04/09/2020	18:11
5	CCV1	1	Water	NA	NA	22.2	2.02	04/09/2020	18:13
6	L2226-01	1	Solid	20.08	20	22.4	6.26	04/09/2020	18:16
7	L2226-01DUP	1	Solid	20.05	20	22.4	6.28	04/09/2020	18:19
8	CCV2	1	Water	NA	NA	22.2	12.02	04/09/2020	18:21