

LB108655


ISM02.4

INSTRUMENT ID : CV1

Linear

 μ Abs.:

41585



A= 0.0000e+000

B= 2.4076e-004

C= 1.1944e-002

Rho= 0.9999809

Accept=Accepted

*slope
& intercept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.016	0.016	18	0.000	18					11
0.05	0.050					149					8
0.20	0.200	0.217	0.017	851	0.0 %	851					11
2.50	2.500	2.485	-0.015	10271	0.0 %	10271					11
5.00	5.000	4.959	-0.041	20546	0.0 %	20546					0
7.50	7.500	7.499	-0.001	31096	0.0 %	31096					0
10.0	10.000	10.025	0.025	41589	0.0 %	41589					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	18	-		0 ISM02.4	PPB	4/10/2020 12:15	Std
	0.2 S0.2	851	-		0.2 ISM02.4	PPB	4/10/2020 12:17	Std
	2.5 S2.5	10271	-		2.5 ISM02.4	PPB	4/10/2020 12:19	Std
	5 S5.0	20546	-		5 ISM02.4	PPB	4/10/2020 12:22	Std
	7.5 S7.5	31096	-		7.5 ISM02.4	PPB	4/10/2020 12:24	Std
	10 S10.0	41589	-		10 ISM02.4	PPB	4/10/2020 12:26	Std
ICV	ICV	15736	3.8005 -		ISM02.4	PPB	4/10/2020 12:29	SMPL
ICB	ICB	-90	-0.0097 -		ISM02.4	PPB	4/10/2020 12:31	SMPL
CCV076	CCV076	20780	5.0149 -		ISM02.4	PPB	4/10/2020 12:34	SMPL
CCB076	CCB076	-106	-0.0136 -		ISM02.4	PPB	4/10/2020 12:36	SMPL
PB128129BL	PBS129	-3	0.0112 -		ISM02.4	PPB	4/10/2020 12:38	SMPL
L2090-01	MCOAR8	2404	0.5907 -		ISM02.4	PPB	4/10/2020 12:40	SMPL
L2090-02	MCOAS1	3136	0.767 -		ISM02.4	PPB	4/10/2020 12:43	SMPL
L2090-03	MCOAS2	460	0.1227 -		ISM02.4	PPB	4/10/2020 12:45	SMPL
L2090-04	MCOAS3	3125	0.7643 -		ISM02.4	PPB	4/10/2020 12:47	SMPL
L2090-05	MCOAS4	456	0.1217 -		ISM02.4	PPB	4/10/2020 12:49	SMPL
L2090-06	MCOAS5	178	0.0548 -		ISM02.4	PPB	4/10/2020 12:52	SMPL
L2090-07	MCOAS6	2365	0.5813 -		ISM02.4	PPB	4/10/2020 12:54	SMPL
L2090-08	MCOAS7	40	0.0216 -		ISM02.4	PPB	4/10/2020 12:56	SMPL
L2090-09	MCOAS7D	54	0.0249 -		ISM02.4	PPB	4/10/2020 12:58	SMPL
L2090-10	MCOAS7S	12007	2.9027 -		ISM02.4	PPB	4/10/2020 13:01	SMPL
CCV077	CCV077	20711	4.9983 -		ISM02.4	PPB	4/10/2020 13:03	SMPL
CCB077	CCB077	-104	-0.0131 -		ISM02.4	PPB	4/10/2020 13:05	SMPL