



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108658

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/10/2020	13:10
2	CAL2	1	Water	NA	NA	20.2	6.99	04/10/2020	13:11
3	CAL3	1	Water	NA	NA	20.2	10.03	04/10/2020	13:12
4	ICV	1	Water	NA	NA	20.1	7.00	04/10/2020	13:15
5	CCV1	1	Water	NA	NA	20.2	2.02	04/10/2020	13:16
6	L2232-01	1	Water	NA	NA	19.5	5.47	04/10/2020	13:19
7	L2232-01DUP	1	Water	NA	NA	19.6	5.48	04/10/2020	13:20
8	CCV2	1	Water	NA	NA	20.1	12.02	04/10/2020	13:22

28108658

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2232 WorkList ID : 137783 Department : Wet-Chemistry Date : 04-10-2020 12:55:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/09/2020	Water	L2232-01	pH	Cool 4 deg C	STER12	H22	CELL-POOL-WATER-2ND-Q	04/09/2020	SM 4500-H B

Date/Time 04.10.2020 13.00
Received by: J0
Relinquished by: J0

Date/Time 04.10.2020 14.00
Received by: J0
Relinquished by: J0