

LB108659


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

54280



A= 0.0000e+000

B= 1.8187e-004

C= 7.9499e-003

Rho= 0.9996429

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	47	0.000	47				
0.2	0.200	0.214	0.014	1135	0.0 %	1135				
2.5	2.500	2.410	-0.090	13210	0.0 %	13210				
5.0	5.000	4.991	-0.009	27398	0.0 %	27398				
7.5	7.500	7.688	0.188	42228	0.0 %	42228				
10.0	10.000	9.880	-0.120	54280	0.0 %	54280				

LB108659 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Apr 2020 14:15:18

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	47	-	0.00007471B	PPB	10 Apr 2020 14:16:23	S	Std
0.2 - 1	50.2	1135	-	0.20007471B	PPB	10 Apr 2020 14:18:39	S	Std
2.5 - 1	52.5	13210	-	2.50007471B	PPB	10 Apr 2020 14:20:57	S	Std
5.0 - 1	55	27398	-	5.00007471B	PPB	10 Apr 2020 14:23:14	S	Std
7.5 - 1	57.5	42228	-	7.50007471B	PPB	10 Apr 2020 14:25:30	S	Std
10.0 - 1	510	54280	-	10.00007471B	PPB	10 Apr 2020 14:27:47	S	Std
ICV90 - 1	ICV90	21950	4.0000	-7471B	PPB	10 Apr 2020 14:30:36	U	SMPL
ICB90 - 1	ICB90	-111	-0.0122	-7471B	PPB	10 Apr 2020 14:32:51	U	SMPL
CCV78 - 1	CCV78	27250	4.9639	-7471B	PPB	10 Apr 2020 14:35:07	U	SMPL
CCB78 - 1	CCB78	-110	-0.0121	-7471B	PPB	10 Apr 2020 14:37:22	U	SMPL
CRA - 1	CRA	1125	0.2126	-7471B	PPB	10 Apr 2020 14:39:38	U	SMPL
HighStd - 1	HighStd	55028	10.0159	-7471B	PPB	10 Apr 2020 14:41:53	U	SMPL
ChkStd - 1	ChkStd	38757	7.0567	-7471B	PPB	10 Apr 2020 14:44:08	U	SMPL
PB128130BL - 1	PBS	-128	-0.0153	-7471B	PPB	10 Apr 2020 14:46:26	U	SMPL
PB128130BS - 1	LCSS	19347	3.5266	-7471B	PPB	10 Apr 2020 14:48:43	U	SMPL
L2180-01 - 1	NEWPORT-TP-7FT	69316	12.6145	-7471B	PPB	10 Apr 2020 14:51:00	U	SMPL
L2195-01 - 1	VNJ-223	352	0.0720	-7471B	PPB	10 Apr 2020 14:53:17	U	SMPL
L2195-05 - 1	VNJ-204	636	0.1236	-7471B	PPB	10 Apr 2020 14:55:36	U	SMPL
L2199-05 - 1	TP-39-38-COMP	1271	0.2391	-7471B	PPB	10 Apr 2020 14:57:53	U	SMPL
L2199-12 - 1	TP-37-36-COMP	26111	4.7568	-7471B	PPB	10 Apr 2020 15:00:08	U	SMPL
CCV79 - 1	CCV79	26897	4.8997	-7471B	PPB	10 Apr 2020 15:02:24	U	SMPL
CCB79 - 1	CCB79	-66	-0.0041	-7471B	PPB	10 Apr 2020 15:04:41	U	SMPL
L2201-01 - 1	CG-SUB	1090	0.2062	-7471B	PPB	10 Apr 2020 15:06:57	U	SMPL
L2205-01 - 1	JC-02-04082020-A	5885	1.0783	-7471B	PPB	10 Apr 2020 15:09:12	U	SMPL
L2205-09 - 1	JC-02-04082020-B	3511	0.6465	-7471B	PPB	10 Apr 2020 15:11:27	U	SMPL
L2205-17 - 1	JC-02-04082020-C	6345	1.1619	-7471B	PPB	10 Apr 2020 15:13:43	U	SMPL
L2205-17DUP - 1	JC-02-04082020-CDUP	6214	1.1381	-7471B	PPB	10 Apr 2020 15:15:59	U	SMPL
L2205-17MS - 1	JC-02-04082020-CMS	25902	4.7188	-7471B	PPB	10 Apr 2020 15:18:15	U	SMPL
L2205-17MSD - 1	JC-02-04082020-CMSD	26896	4.8995	-7471B	PPB	10 Apr 2020 15:20:31	U	SMPL
PB128131BL - 1	PBS	-80	-0.0066	-7471B	PPB	10 Apr 2020 15:22:49	U	SMPL
PB128131BS - 1	LCSS	20038	3.6523	-7471B	PPB	10 Apr 2020 15:25:07	U	SMPL
L2215-02 - 1	TP-9	181	0.0409	-7471B	PPB	10 Apr 2020 15:27:23	U	SMPL
CCV80 - 1	CCV80	27591	5.0259	-7471B	PPB	10 Apr 2020 15:29:38	U	SMPL
CCB80 - 1	CCB80	-71	-0.0050	-7471B	PPB	10 Apr 2020 15:31:53	U	SMPL
L2215-02DUP - 1	TP-9DUP	391	0.0791	-7471B	PPB	10 Apr 2020 15:34:10	U	SMPL
L2215-02MS - 1	TP-9MS	20538	3.7432	-7471B	PPB	10 Apr 2020 15:36:26	U	SMPL
L2215-02MSD - 1	TP-9MSD	20477	3.7321	-7471B	PPB	10 Apr 2020 15:38:41	U	SMPL
L2215-05 - 1	TP-10	1060	0.2007	-7471B	PPB	10 Apr 2020 15:40:58	U	SMPL
L2215-08 - 1	TP-8	844	0.1614	-7471B	PPB	10 Apr 2020 15:43:14	U	SMPL
L2215-11 - 1	TP-6	209	0.0460	-7471B	PPB	10 Apr 2020 15:45:30	U	SMPL
L2215-14 - 1	TP-4	258	0.0549	-7471B	PPB	10 Apr 2020 15:47:46	U	SMPL
L2215-17 - 1	TP-3	361	0.0736	-7471B	PPB	10 Apr 2020 15:50:03	U	SMPL
L2215-20 - 1	TP-5	240	0.0516	-7471B	PPB	10 Apr 2020 15:56:16	U	SMPL
L2215-23 - 1	TP-7	890	0.1698	-7471B	PPB	10 Apr 2020 15:58:31	U	SMPL
CCV81 - 1	CCV81	27565	5.0212	-7471B	PPB	10 Apr 2020 16:00:47	U	SMPL
CCB81 - 1	CCB81	-77	-0.0061	-7471B	PPB	10 Apr 2020 16:03:02	U	SMPL
L2215-26 - 1	TP-1	136	0.0327	-7471B	PPB	10 Apr 2020 16:05:19	U	SMPL
L2215-29 - 1	TP-2	499	0.0987	-7471B	PPB	10 Apr 2020 16:07:34	U	SMPL
L2222-01 - 1	S-1	3309	0.6098	-7471B	PPB	10 Apr 2020 16:09:50	U	SMPL
L2223-01 - 1	FK-BPM-03-015-ABCD	1300	0.2444	-7471B	PPB	10 Apr 2020 16:12:05	U	SMPL
L2224-01 - 1	FK-BPM-04-009-ABCD	963	0.1831	-7471B	PPB	10 Apr 2020 16:14:21	U	SMPL
L2225-01 - 1	FK-BPM-01-C3-D3-C3A-D3A	1776	0.3310	-7471B	PPB	10 Apr 2020 16:16:37	U	SMPL
L2226-01 - 1	FK-PS-01-066	2406	0.4455	-7471B	PPB	10 Apr 2020 16:18:54	U	SMPL
CCV82 - 1	CCV82	27370	4.9857	-7471B	PPB	10 Apr 2020 16:21:10	U	SMPL
CCB82 - 1	CCB82	-90	-0.0084	-7471B	PPB	10 Apr 2020 16:23:27	U	SMPL
L2205-17LX5 - 1		1210	0.2280	-7471B	PPB	10 Apr 2020 16:25:43	U	SMPL
L2205-17A - 1		27865	5.0758	-7471B	PPB	10 Apr 2020 16:27:58	U	SMPL
L2215-02LX5 - 1		-72	-0.0051	-7471B	PPB	10 Apr 2020 16:30:14	U	SMPL
L2215-02A - 1		21977	4.0049	-7471B	PPB	10 Apr 2020 16:32:30	U	SMPL
L2180-01DLX2 - 1		34596	6.2999	-7471B	PPB	10 Apr 2020 16:34:46	U	SMPL
CCV83 - 1	CCV83	27484	5.0065	-7471B	PPB	10 Apr 2020 16:37:02	U	SMPL
CCB83 - 1	CCB83	-67	-0.0042	-7471B	PPB	10 Apr 2020 16:39:19	U	SMPL