



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 4/12/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 04/10/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:00  
Out Date: 04/11/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB108667

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2195-01	VNJ-223	1	1.13	9.67	9.15	93.9	
L2195-05	VNJ-204	2	1.13	9.56	9.07	94.2	
L2229-13	TP-32	3	1.16	9.52	8.14	83.5	
L2229-14	TP-32-EPH-2	4	1.16	9.52	8.14	83.5	
L2229-15	TP-32	5	1.15	9.61	8.21	83.5	
L2229-37	TP-10	6	1.14	9.61	7.64	76.7	
L2229-38	TP-10-EPH-2	7	1.14	9.61	7.64	76.7	
L2229-39	TP-10	8	1.14	9.98	8.35	81.6	
L2230-13	TP-18	9	1.14	9.88	8.18	80.5	
L2230-14	TP-18-EPH-2	10	1.14	9.88	8.18	80.5	
L2230-15	TP-18	11	1.15	9.5	8.11	83.4	
L2230-17	TP-12-A	12	1.16	9.56	8.00	81.4	
L2230-18	TP-12-A-EPH-2	13	1.16	9.56	8.00	81.4	
L2230-19	TP-12-A	14	1.14	9.56	8.02	81.7	
L2230-21	TP-12-B	15	1.15	9.85	7.67	74.9	
L2230-22	TP-12-B-EPH-2	16	1.15	9.85	7.67	74.9	
L2230-23	TP-12-B	17	1.13	9.98	8.8	86.7	
L2230-29	TP-20	18	1.14	9.96	8.14	79.4	
L2230-30	TP-20-EPH-2	19	1.14	9.96	8.14	79.4	
L2230-31	TP-20	20	1.13	9.84	8.01	79.0	
L2230-37	TP-33	21	1.15	9.87	7.69	75.0	
L2230-38	TP-33-EPH-2	22	1.15	9.87	7.69	75.0	
L2230-39	TP-33	23	1.15	9.9	7.69	74.7	
L2230-45	TP-34	24	1.16	9.78	8.18	81.4	
L2230-46	TP-34-EPH-2	25	1.16	9.78	8.18	81.4	
L2230-47	TP-34	26	1.16	9.98	8.34	81.4	
L2230-49	CONCRETE-4-9-2020	27	1.00	2.00	2.00	100.0	CONCRETE SAMPLE 100% SOLIDS
L2230-51	TP-1-9-17	28	1.16	9.63	7.86	79.1	



PERCENT SOLID

Supervisor: apatel  
Analyst: JIGNESH  
Date: 4/12/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:25  
In Date: 04/10/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:00  
Out Date: 04/11/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB108667

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2230-52	TP-1-9-17-EPH-2	29	1.16	9.63	7.86	79.1	
L2230-55	TP-2-3-4	30	1.15	9.56	7.73	78.2	
L2230-56	TP-2-3-4-EPH-2	31	1.15	9.56	7.73	78.2	
L2230-59	TP-11-19	32	1.15	9.97	8.22	80.2	
L2230-60	TP-11-19-EPH-2	33	1.15	9.97	8.22	80.2	
L2230-63	TP-25-31-39	34	1.16	9.6	8.02	81.3	
L2230-64	TP-25-31-39-EPH-2	35	1.16	9.6	8.02	81.3	
L2230-67	TP-40-41-42	36	1.16	9.8	8.31	82.8	
L2230-68	TP-40-41-42-EPH-2	37	1.16	9.8	8.31	82.8	
L2231-05	SVOC-GPC-BLANK	38	1.00	2.00	2.00	100.0	
L2231-06	PEST-GPC-BLANK	39	1.00	2.00	2.00	100.0	
L2231-07	PEST-GPC-BLANK-SPIKE	40	1.00	2.00	2.00	100.0	
L2231-08	PCB-GPC-BLANK	41	1.00	2.00	2.00	100.0	
L2231-09	PCB-GPC-BLANK-SPIKE	42	1.00	2.00	2.00	100.0	
L2231-10	SVOC-GPC2-BLANK	43	1.00	2.00	2.00	100.0	
L2231-11	PEST-GCP2-BLANK	44	1.00	2.00	2.00	100.0	
L2231-12	PEST-GCP2-BLANK-SPIKE	45	1.00	2.00	2.00	100.0	
L2231-13	PCB-GCP2-BLANK	46	1.00	2.00	2.00	100.0	
L2231-14	PCB-GCP2-BLANK-SPIKE	47	1.00	2.00	2.00	100.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$