



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 4/14/2020

OVENTEMP IN Celsius(°C): 106
Time IN: 15:10
In Date: 04/13/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:34
Out Date: 04/14/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108683

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L2236-01	METXN7	1	1.15	9.75	7.83	77.7	
L2236-02	METXN8	2	1.15	9.96	7.93	77.0	
L2236-03	METXN9	3	1.16	9.75	8.24	82.4	
L2236-04	METXP0	4	1.19	9.73	7.88	78.3	
L2236-05	METXP1	5	1.14	9.97	7.81	75.5	
L2236-06	METXP2	6	1.13	9.86	8.17	80.6	
L2236-07	METXP3	7	1.18	9.64	7.96	80.1	
L2236-08	METXP4	8	1.17	9.96	7.9	76.6	
L2236-09	METXP5	9	1.18	9.66	7.42	73.6	
L2236-10	METXP6	10	1.18	9.51	7.67	77.9	
L2236-11	METXP6D	11	1.18	9.51	7.67	77.9	
L2236-12	METXP6S	12	1.18	9.51	7.67	77.9	
L2236-13	METXP7	13	1.14	9.73	8.39	84.4	
L2236-14	METXP8	14	1.17	9.58	8.26	84.3	
L2236-15	METXP9	15	1.14	9.81	7.92	78.2	
L2236-16	METXQ1	16	1.18	9.66	8.53	86.7	
L2236-17	METXQ2	17	1.11	9.52	8.19	84.2	
L2236-18	METXQ3	18	1.19	9.72	8.06	80.5	
L2236-19	METXQ4	19	1.15	9.94	8.32	81.6	
L2236-20	METXQ5	20	1.14	9.8	8.4	83.8	
L2236-21	METXQ6	21	1.12	9.67	8.62	87.7	
L2236-22	METXQ7	22	1.12	9.93	8.33	81.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$