



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: Daniel

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108686

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/13/2020	09:00
2	L2171-02		1	8	100.00	>212.0	>212.0	04/13/2020	09:20
3	L2171-02DUP		1	9	100.00	>212.0	>212.0	04/13/2020	09:45
4	L2185-01		1	8	100.00	>212.0	>212.0	04/13/2020	10:15
5	L2185-16		1	8	100.00	>212.0	>212.0	04/13/2020	10:35
6	L2186-11		1	8	100.00	>212.0	>212.0	04/13/2020	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LN 128686

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp041320 WorkList ID : 137832 Department : Wet-Chemistry Date : 04-13-2020 08:33:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/13/2020	Water	L2171-02	Flash Point	Cool 4 deg C	PSEG01	H11	FRAC-COMP	04/06/2020	1010A
04/14/2020	Water	L2185-01	Flash Point	Cool 4 deg C	PSEG01	H12	30773	04/07/2020	1010A
04/14/2020	Water	L2185-16	Flash Point	Cool 4 deg C	PSEG01	H12	30928-2:1-COMP	04/07/2020	1010A
04/14/2020	Water	L2186-11	Flash Point	Cool 4 deg C	PSEG01	H12	COMP-7:1-WATER	04/07/2020	1010A

Date/Time 04/13/2020 8:40
 Received by: JK
 Relinquished by: JP

Date/Time 04/13/2020 11:15
 Received by: JP
 Relinquished by: JK