



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108691

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/13/2020	16:00
2	CAL2	1	Water	NA	NA	20.2	7.01	04/13/2020	16:02
3	CAL3	1	Water	NA	NA	20.2	10.04	04/13/2020	16:05
4	ICV	1	Water	NA	NA	20.1	7.00	04/13/2020	16:06
5	CCV1	1	Water	NA	NA	20.2	2.01	04/13/2020	16:10
6	L2256-01	1	Water	NA	NA	20.4	5.31	04/13/2020	16:11
7	L2256-01DUP	1	Water	NA	NA	20.5	5.32	04/13/2020	16:12
8	L2256-02	1	Water	NA	NA	20.5	5.08	04/13/2020	16:14
9	L2270-01	1	Water	NA	NA	21.2	4.91	04/13/2020	16:15
10	CCV2	1	Water	NA	NA	20.1	12.02	04/13/2020	16:16

108691

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2256 WorkList ID : 137833 Department : Wet-Chemistry Date : 04-13-2020 15:45:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/10/2020	Water	L2256-01 - <i>1x</i>	pH	Cool 4 deg C	STER16	G41	MAIN-CELL	04/10/2020	SM 4500-H B
04/10/2020	Water	L2256-02 - <i>1x</i>	pH	Cool 4 deg C	STER16	G41	EX-CELL	04/10/2020	SM 4500-H B
04/09/2020	Water	L2270-01 - <i>1x</i>	pH	Cool 4 deg C	STER14	G61	MINI-CELL	04/09/2020	SM 4500-H B

Date/Time 04/13/2020 15:50
 Received by: *Ja*
 Relinquished by: *Ja*

Date/Time 04-13-2020 17:100
 Received by: *Ja*
 Relinquished by: *Ja*