


LB108696

7470A

INSTRUMENT ID: CV1

 μ Abs.:

40947



Linear

A= 0.0000e+000

B= 2.4734e-004

C= -2.5962e-002

Rho= 0.9997479

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	74	0.000	74				
0.2	0.200	0.221	0.021	1000	0.0 %	1000				
2.5	2.500	2.525	0.025	10314	0.0 %	10314				
5.0	5.000	5.029	0.029	20436	0.0 %	20436				
7.5	7.500	7.330	-0.170	29741	0.0 %	29741				
10.0	10.000	10.102	0.102	40947	0.0 %	40947				

LB108696 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Apr 2020 10:30:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	74	-	0.0000	7470A	PPB	14 Apr 2020 10:37:07	S	Std
0.2 - 1	50.2	1000	-	0.2000	7470A	PPB	14 Apr 2020 10:39:15	S	Std
2.5 - 1	52.5	10314	-	2.5000	7470A	PPB	14 Apr 2020 10:41:23	S	Std
5.0 - 1	55	20436	-	5.0000	7470A	PPB	14 Apr 2020 10:43:30	S	Std
7.5 - 1	57.5	29741	-	7.5000	7470A	PPB	14 Apr 2020 10:45:37	S	Std
10.0 - 1	510	40947	-	10.0000	7470A	PPB	14 Apr 2020 10:47:43	S	Std
ICV93 - 1	ICV93	15042	3.6946	-7470A	PPB	14 Apr 2020 10:50:19	U	SMPL	
ICB93 - 1	ICB93	-48	-0.0378	-7470A	PPB	14 Apr 2020 10:52:24	U	SMPL	
CCV97 - 1	CCV97	20365	5.0112	-7470A	PPB	14 Apr 2020 10:54:30	U	SMPL	
CCB97 - 1	CCB97	-68	-0.0428	-7470A	PPB	14 Apr 2020 10:56:36	U	SMPL	
CRA - 1	CRA	959	0.2112	-7470A	PPB	14 Apr 2020 10:58:42	U	SMPL	
HighStd - 1	HighStd	40034	9.8762	-7470A	PPB	14 Apr 2020 11:00:48	U	SMPL	
ChkStd - 1	ChkStd	27549	6.7881	-7470A	PPB	14 Apr 2020 11:02:54	U	SMPL	
PB128185BL - 1	PBW	252	0.0364	-7470A	PPB	14 Apr 2020 11:05:02	U	SMPL	
PB128185BS - 1	LCSW	14371	3.5286	-7470A	PPB	14 Apr 2020 11:07:09	U	SMPL	
L2229-16 - 1	TP-32	264	0.0393	-7470A	PPB	14 Apr 2020 11:09:16	U	SMPL	
L2229-16DUP - 1	TP-32DUP	324	0.0542	-7470A	PPB	14 Apr 2020 11:14:08	U	SMPL	
L2229-16MS - 1	TP-32MS	13801	3.3876	-7470A	PPB	14 Apr 2020 11:16:15	U	SMPL	
L2229-16MSD - 1	TP-32MSD	13962	3.4275	-7470A	PPB	14 Apr 2020 11:18:23	U	SMPL	
L2229-40 - 1	TP-10	781	0.1672	-7470A	PPB	14 Apr 2020 11:20:29	U	SMPL	
CCV98 - 1	CCV98	20400	5.0199	-7470A	PPB	14 Apr 2020 11:22:35	U	SMPL	
CCB98 - 1	CCB98	-43	-0.0366	-7470A	PPB	14 Apr 2020 11:24:41	U	SMPL	
L2230-16 - 1	TP-18	800	0.1719	-7470A	PPB	14 Apr 2020 11:26:47	U	SMPL	
L2230-20 - 1	TP-12-A	987	0.2182	-7470A	PPB	14 Apr 2020 11:28:52	U	SMPL	
L2230-24 - 1	TP-12-B	396	0.0720	-7470A	PPB	14 Apr 2020 11:30:58	U	SMPL	
L2230-32 - 1	TP-20	15	-0.0223	-7470A	PPB	14 Apr 2020 11:33:04	U	SMPL	
L2230-40 - 1	TP-33	478	0.0923	-7470A	PPB	14 Apr 2020 11:35:10	U	SMPL	
L2230-48 - 1	TP-34	490	0.0952	-7470A	PPB	14 Apr 2020 11:37:17	U	SMPL	
L2230-54 - 1	TP-1-9-17	509	0.0999	-7470A	PPB	14 Apr 2020 11:39:25	U	SMPL	
L2230-58 - 1	TP-2-3-4	614	0.1259	-7470A	PPB	14 Apr 2020 11:41:32	U	SMPL	
L2230-62 - 1	TP-11-19	554	0.1111	-7470A	PPB	14 Apr 2020 11:43:40	U	SMPL	
L2230-66 - 1	TP-25-31-39	576	0.1165	-7470A	PPB	14 Apr 2020 11:45:47	U	SMPL	
CCV99 - 1	CCV99	20245	4.9815	-7470A	PPB	14 Apr 2020 11:47:53	U	SMPL	
CCB99 - 1	CCB99	5	-0.0247	-7470A	PPB	14 Apr 2020 11:49:59	U	SMPL	
L2230-70 - 1	TP-40-41-42	378	0.0675	-7470A	PPB	14 Apr 2020 11:52:06	U	SMPL	
L2242-04 - 1	B-1	556	0.1116	-7470A	PPB	14 Apr 2020 11:54:12	U	SMPL	
L2243-06 - 1	TP-34-33-COMP	344	0.0591	-7470A	PPB	14 Apr 2020 11:56:18	U	SMPL	
L2243-13 - 1	TP-31-30-COMP	382	0.0685	-7470A	PPB	14 Apr 2020 11:58:25	U	SMPL	
L2243-20 - 1	TP-28-COMP	377	0.0673	-7470A	PPB	14 Apr 2020 12:00:31	U	SMPL	
L2244-06 - 1	TP-25-24-COMP	367	0.0648	-7470A	PPB	14 Apr 2020 12:02:38	U	SMPL	
L2244-13 - 1	TP-23-22-COMP	373	0.0663	-7470A	PPB	14 Apr 2020 12:04:46	U	SMPL	
PB128185TB - 1	PB128160TB	422	0.0784	-7470A	PPB	14 Apr 2020 12:06:53	U	SMPL	
L2229-16LX5 - 1		126	0.0052	-7470A	PPB	14 Apr 2020 12:09:01	U	SMPL	
L2229-16A - 1		14696	3.6090	-7470A	PPB	14 Apr 2020 12:11:08	U	SMPL	
CCV01 - 1	CCV01	19840	4.8814	-7470A	PPB	14 Apr 2020 12:13:14	U	SMPL	
CCB01 - 1	CCB01	-41	-0.0361	-7470A	PPB	14 Apr 2020 12:15:20	U	SMPL	