


LB108701

7471B

INSTRUMENT ID: CU1

 μ Abs.:

57057



Linear

A= 0.0000e+000

B= 1.7634e-004

C= -6.5485e-002

Rho= 0.9999898

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	244	0.000	244				
0.2	0.200	0.211	0.011	1566	0.0 %	1566				
2.5	2.500	2.502	0.002	14561	0.0 %	14561				
5.0	5.000	5.028	0.028	28887	0.0 %	28887				
7.5	7.500	7.485	-0.015	42819	0.0 %	42819				
10.0	10.000	9.996	-0.004	57057	0.0 %	57057				

LB108701 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 14 Apr 2020 12:55:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	244	-	0.00007471B	PPB		14 Apr 2020 13:30:13	S	Std
0.2 - 1	50.2	1566	-	0.20007471B	PPB		14 Apr 2020 13:32:30	S	Std
2.5 - 1	52.5	14561	-	2.50007471B	PPB		14 Apr 2020 13:34:47	S	Std
5.0 - 1	55	28887	-	5.00007471B	PPB		14 Apr 2020 13:37:04	S	Std
7.5 - 1	57.5	42819	-	7.50007471B	PPB		14 Apr 2020 13:39:20	S	Std
10.0 - 1	510	57057	-	10.00007471B	PPB		14 Apr 2020 13:41:37	S	Std
ICV94 - 1	ICV94	21164	3.6666	-7471B	PPB		14 Apr 2020 13:44:21	U	SMPL
ICB94 - 1	ICB94	155	-0.0382	-7471B	PPB		14 Apr 2020 13:46:35	U	SMPL
CCV02 - 1	CCV02	29176	5.0794	-7471B	PPB		14 Apr 2020 13:48:52	U	SMPL
CCB02 - 1	CCB02	141	-0.0406	-7471B	PPB		14 Apr 2020 13:51:06	U	SMPL
CRA - 1	CRA	1534	0.2050	-7471B	PPB		14 Apr 2020 13:53:22	U	SMPL
HighStd - 1	HighStd	56430	9.8854	-7471B	PPB		14 Apr 2020 13:55:37	U	SMPL
ChkStd - 1	ChkStd	40537	7.0828	-7471B	PPB		14 Apr 2020 13:57:52	U	SMPL
PB128207BL - 1	PBS	139	-0.0410	-7471B	PPB		14 Apr 2020 14:00:10	U	SMPL
PB128207BS - 1	LCSS	21195	3.6720	-7471B	PPB		14 Apr 2020 14:02:27	U	SMPL
L2115-01 - 1	CL-01-041020	2525	0.3798	-7471B	PPB		14 Apr 2020 14:04:43	U	SMPL
L2115-01DUP - 1	CL-01-041020DUP	2629	0.3981	-7471B	PPB		14 Apr 2020 14:06:59	U	SMPL
L2115-01MS - 1	CL-01-041020MS	23934	4.1550	-7471B	PPB		14 Apr 2020 14:09:16	U	SMPL
L2115-01MSD - 1	CL-01-041020MSD	23977	4.1626	-7471B	PPB		14 Apr 2020 14:11:33	U	SMPL
L2198-05 - 1	TP-43-COMP	958	0.1034	-7471B	PPB		14 Apr 2020 14:13:48	U	SMPL
CCV03 - 1	CCV03	29144	5.0738	-7471B	PPB		14 Apr 2020 14:16:05	U	SMPL
CCB03 - 1	CCB03	193	-0.0315	-7471B	PPB		14 Apr 2020 14:18:20	U	SMPL
L2198-12 - 1	TP-42-41-COMP	1758	0.2445	-7471B	PPB		14 Apr 2020 14:20:35	U	SMPL
L2245-06 - 1	TP-26	5578	0.9181	-7471B	PPB		14 Apr 2020 14:22:50	U	SMPL
L2245-10 - 1	TP-23B	191435	33.6921	-7471B	PPB		14 Apr 2020 14:25:06	U	SMPL
L2245-14 - 1	TP-23A	128237	22.5478	-7471B	PPB		14 Apr 2020 14:27:22	U	SMPL
L2245-18 - 1	TP-22	73707	12.9320	-7471B	PPB		14 Apr 2020 14:29:42	U	SMPL
L2245-22 - 1	TP-21A	74976	13.1558	-7471B	PPB		14 Apr 2020 14:32:01	U	SMPL
L2245-26 - 1	TP-21B	66430	11.6488	-7471B	PPB		14 Apr 2020 14:34:20	U	SMPL
L2245-30 - 1	TP-28	31389	5.4696	-7471B	PPB		14 Apr 2020 14:36:39	U	SMPL
L2245-34 - 1	TP-27	19883	3.4407	-7471B	PPB		14 Apr 2020 14:38:59	U	SMPL
L2245-38 - 1	TP-5-13-14	73560	12.9061	-7471B	PPB		14 Apr 2020 14:41:16	U	SMPL
CCV04 - 1	CCV04	29013	5.0507	-7471B	PPB		14 Apr 2020 14:43:31	U	SMPL
CCB04 - 1	CCB04	167	-0.0360	-7471B	PPB		14 Apr 2020 14:45:49	U	SMPL
L2245-41 - 1	TP-7-15	63614	11.1522	-7471B	PPB		14 Apr 2020 14:48:04	U	SMPL
L2255-05 - 1	TP-27-26-COMP	1734	0.2403	-7471B	PPB		14 Apr 2020 14:50:20	U	SMPL
L2115-01LX5 - 1		539	0.0296	-7471B	PPB		14 Apr 2020 14:52:38	U	SMPL
L2115-01A - 1		24043	4.1743	-7471B	PPB		14 Apr 2020 14:54:53	U	SMPL
L2245-10DLX10 - 1		20331	3.5197	-7471B	PPB		14 Apr 2020 14:57:09	U	SMPL
L2245-14DLX5 - 1		26462	4.6008	-7471B	PPB		14 Apr 2020 14:59:25	U	SMPL
L2245-18DLX2 - 1		37417	6.5326	-7471B	PPB		14 Apr 2020 15:01:41	U	SMPL
L2245-22DLX2 - 1		37997	6.6349	-7471B	PPB		14 Apr 2020 15:03:58	U	SMPL
L2245-26DLX2 - 1		33324	5.8109	-7471B	PPB		14 Apr 2020 15:06:16	U	SMPL
L2245-38DLX2 - 1		36683	6.4032	-7471B	PPB		14 Apr 2020 15:08:34	U	SMPL
CCV05 - 1	CCV05	29005	5.0493	-7471B	PPB		14 Apr 2020 15:10:50	U	SMPL
CCB05 - 1	CCB05	173	-0.0350	-7471B	PPB		14 Apr 2020 15:13:07	U	SMPL
L2245-41DLX2 - 1		32529	5.6707	-7471B	PPB		14 Apr 2020 15:15:23	U	SMPL
CCV06 - 1	CCV06	29259	5.0940	-7471B	PPB		14 Apr 2020 15:17:39	U	SMPL
CCB06 - 1	CCB06	209	-0.0286	-7471B	PPB		14 Apr 2020 15:19:56	U	SMPL