



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108709

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/14/2020	17:05
2	CAL2	1	Water	NA	NA	20.2	7.01	04/14/2020	17:06
3	CAL3	1	Water	NA	NA	20.2	10.03	04/14/2020	17:08
4	ICV	1	Water	NA	NA	20.2	6.99	04/14/2020	17:10
5	CCV1	1	Water	NA	NA	20.2	2.01	04/14/2020	17:11
6	L2123-09	1	Solid	20.02	20	21.8	8.71	04/14/2020	17:12
7	L2252-01	1	Solid	20.03	20	21.9	6.71	04/14/2020	17:15
8	L2252-02	1	Solid	20.01	20	21.7	6.95	04/14/2020	17:16
9	L2259-01	1	Solid	20.03	20	21.7	7.16	04/14/2020	17:19
10	L2259-02	1	Solid	20.04	20	21.1	6.72	04/14/2020	17:20
11	L2259-03	1	Solid	20.02	20	21.6	7.39	04/14/2020	17:22
12	L2277-04	1	Solid	20.03	20	22.1	8.17	04/14/2020	17:25
13	L2278-01	1	Solid	20.04	20	21.4	7.77	04/14/2020	17:30
14	L2278-02	1	Solid	20.02	20	21.0	7.04	04/14/2020	17:32
15	L2278-03	1	Solid	20.03	20	20.9	6.58	04/14/2020	17:35
16	CCV2	1	Water	NA	NA	20.1	12.02	04/14/2020	17:36
17	L2281-11	1	Solid	20.04	20	21.4	7.97	04/14/2020	17:39
18	L2281-14	1	Solid	20.02	20	20.9	8.17	04/14/2020	17:42
19	L2281-14DUP	1	Solid	20.03	20	20.8	8.18	04/14/2020	17:43
20	CCV3	1	Water	NA	NA	20.1	2.01	04/14/2020	17:45

B108709

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L2277 WorkList ID : 137875 Department : Wet-Chemistry Date : 04-14-2020 12:32:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/14/2020	Solid	L2123-09	Corrosivity	Cool 4 deg C	PSEG05		TR-04-142004	04/14/2020	9045D
04/10/2020	Solid	L2252-01	Corrosivity	Cool 4 deg C	CLEA11	H31	CHASSIS-WASH-SEDIMENT-	04/10/2020	9045D
04/10/2020	Solid	L2252-02	Corrosivity	Cool 4 deg C	CLEA11	H31	BUS-WASH-PIT	04/10/2020	9045D
04/09/2020	Solid	L2259-01	Corrosivity	Cool 4 deg C	CLEA11		CHASSIS-BUS-WASH	04/09/2020	9045D
04/09/2020	Solid	L2259-02	Corrosivity	Cool 4 deg C	CLEA11		OIL-WATER-SEPARATOR-2	04/09/2020	9045D
04/09/2020	Solid	L2259-03	Corrosivity	Cool 4 deg C	CLEA11		BUS-WASH	04/09/2020	9045D
04/14/2020	Solid	L2277-04	Corrosivity	Cool 4 deg C	PSEG01	H31	B-2	04/14/2020	9045D
04/13/2020	Solid	L2278-01	Corrosivity	Cool 4 deg C	CLEA11		CHASSIS-WASH	04/13/2020	9045D
04/13/2020	Solid	L2278-02	Corrosivity	Cool 4 deg C	CLEA11		BUS-WASH	04/13/2020	9045D
04/13/2020	Solid	L2278-03	Corrosivity	Cool 4 deg C	CLEA11		WASTE-RECLAMATION-PIT	04/13/2020	9045D
04/14/2020	Solid	L2281-11	Corrosivity	Cool 4 deg C	PSEG01	H31	TP-35-36-43-44	04/14/2020	9045D
04/14/2020	Solid	L2281-14	Corrosivity	Cool 4 deg C	PSEG01	H31	TP-37-38-45-46	04/14/2020	9045D

Date/Time 04-14-2020 16:00
 Received by: Jr
 Relinquished by: 40

Date/Time 04-14-2020 18:10
 Received by: Jr
 Relinquished by: Jr