


LB108715

7470A

INSTRUMENT ID: CV4

μ Abs.:

37241



Linear

A= 0.0000e+000

B= 2.6342e-004

C= -1.8252e-002

Rho= 0.9993246

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-3	0.000	-3				
0.2	0.200	0.182	-0.018	759	0.0 %	759				
2.5	2.500	2.417	-0.083	9243	0.0 %	9243				
5.0	5.000	5.113	0.113	19481	0.0 %	19481				
7.5	7.500	7.716	0.216	29359	0.0 %	29359				
10.0	10.000	9.792	-0.208	37241	0.0 %	37241				

LB108715 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Apr 2020 07:51:50

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7470A	PPB	15 Apr 2020 07:59:00	S	Std
0.2 - 1	50.2	759	-	0.2000	7470A	PPB	15 Apr 2020 08:01:08	S	Std
2.5 - 1	52.5	9243	-	2.5000	7470A	PPB	15 Apr 2020 08:03:15	S	Std
5.0 - 1	55	19481	-	5.0000	7470A	PPB	15 Apr 2020 08:05:22	S	Std
7.5 - 1	57.5	29359	-	7.5000	7470A	PPB	15 Apr 2020 08:07:30	S	Std
10.0 - 1	510	37241	-	10.0000	7470A	PPB	15 Apr 2020 08:09:36	S	Std
ICV95 - 1	ICV95	14878	3.9009	-7470A	PPB	PPB	15 Apr 2020 08:12:17	U	SMPL
ICB95 - 1	ICB95	-118	-0.0493	-7470A	PPB	PPB	15 Apr 2020 08:14:23	U	SMPL
CCV09 - 1	CCV09	18793	4.9322	-7470A	PPB	PPB	15 Apr 2020 08:16:28	U	SMPL
CCB09 - 1	CCB09	-138	-0.0546	-7470A	PPB	PPB	15 Apr 2020 08:18:34	U	SMPL
CRA - 1	CRA	762	0.1825	-7470A	PPB	PPB	15 Apr 2020 08:20:40	U	SMPL
HighStd - 1	HighStd	37767	9.9304	-7470A	PPB	PPB	15 Apr 2020 08:22:46	U	SMPL
ChkStd - 1	ChkStd	26021	6.8362	-7470A	PPB	PPB	15 Apr 2020 08:24:51	U	SMPL
PB128222BL - 1	PBW	-187	-0.0675	-7470A	PPB	PPB	15 Apr 2020 08:26:59	U	SMPL
PB128222BS - 1	LCSW	14143	3.7073	-7470A	PPB	PPB	15 Apr 2020 08:29:07	U	SMPL
L2115-08 - 1	CL-01-041020	-100	-0.0446	-7470A	PPB	PPB	15 Apr 2020 08:31:14	U	SMPL
L2115-08DUP - 1	CL-01-041020DUP	-4	-0.0193	-7470A	PPB	PPB	15 Apr 2020 08:33:21	U	SMPL
L2115-08MS - 1	CL-01-041020MS	14271	3.7410	-7470A	PPB	PPB	15 Apr 2020 08:35:28	U	SMPL
L2115-08MSD - 1	CL-01-041020MSD	13223	3.4650	-7470A	PPB	PPB	15 Apr 2020 08:37:36	U	SMPL
L2122-10 - 1	SU-04-041020	-77	-0.0385	-7470A	PPB	PPB	15 Apr 2020 08:39:42	U	SMPL
CCV10 - 1	CCV10	19466	5.1095	-7470A	PPB	PPB	15 Apr 2020 08:41:48	U	SMPL
CCB10 - 1	CCB10	-107	-0.0464	-7470A	PPB	PPB	15 Apr 2020 08:43:54	U	SMPL
L2198-06 - 1	TP-43-COMP	8	-0.0161	-7470A	PPB	PPB	15 Apr 2020 08:45:59	U	SMPL
L2198-13 - 1	TP-42-41-COMP	17	-0.0138	-7470A	PPB	PPB	15 Apr 2020 08:48:06	U	SMPL
L2245-09 - 1	TP-26	60	-0.0024	-7470A	PPB	PPB	15 Apr 2020 08:50:12	U	SMPL
L2245-13 - 1	TP-23B	19	-0.0132	-7470A	PPB	PPB	15 Apr 2020 08:52:18	U	SMPL
L2245-17 - 1	TP-23A	32	-0.0098	-7470A	PPB	PPB	15 Apr 2020 08:54:25	U	SMPL
L2245-21 - 1	TP-22	26	-0.0114	-7470A	PPB	PPB	15 Apr 2020 08:56:32	U	SMPL
L2245-25 - 1	TP-21A	50	-0.0051	-7470A	PPB	PPB	15 Apr 2020 08:58:39	U	SMPL
L2245-29 - 1	TP-21B	41	-0.0075	-7470A	PPB	PPB	15 Apr 2020 09:00:47	U	SMPL
L2245-33 - 1	TP-28	46	-0.0061	-7470A	PPB	PPB	15 Apr 2020 09:02:55	U	SMPL
L2245-37 - 1	TP-27	30	-0.0103	-7470A	PPB	PPB	15 Apr 2020 09:05:01	U	SMPL
CCV11 - 1	CCV11	19434	5.1011	-7470A	PPB	PPB	15 Apr 2020 09:07:07	U	SMPL
CCB11 - 1	CCB11	-100	-0.0446	-7470A	PPB	PPB	15 Apr 2020 09:09:14	U	SMPL
L2245-40 - 1	TP-5-13-14	42	-0.0072	-7470A	PPB	PPB	15 Apr 2020 09:11:20	U	SMPL
L2245-43 - 1	TP-7-15	83	0.0036	-7470A	PPB	PPB	15 Apr 2020 09:13:26	U	SMPL
L2255-06 - 1	TP-27-26-COMP	7	-0.0164	-7470A	PPB	PPB	15 Apr 2020 09:15:32	U	SMPL
PB128222TB - 1	PB128174TB	31	-0.0101	-7470A	PPB	PPB	15 Apr 2020 09:17:39	U	SMPL
L2115-08LX5 - 1		42	-0.0072	-7470A	PPB	PPB	15 Apr 2020 09:19:45	U	SMPL
L2115-08A - 1		15001	3.9333	-7470A	PPB	PPB	15 Apr 2020 09:21:52	U	SMPL
CCV12 - 1	CCV12	18763	4.9243	-7470A	PPB	PPB	15 Apr 2020 09:23:59	U	SMPL
CCB12 - 1	CCB12	-98	-0.0441	-7470A	PPB	PPB	15 Apr 2020 09:26:07	U	SMPL