


LB108717

ISM02.4

INSTRUMENT ID : CV1

μ Abs.:

43565



Linear

A= 0.0000e+000

B= 2.3106e-004

C= -7.1061e-003

Rho= 0.9999522

Accept=Accepted

*slope
y-intercept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.013	0.013	85	0.000	85					14
0.05	0.050	0.057	0.007	279	0.0 %	279					1
0.20	0.200	0.201	0.001	899	0.0 %	899					0
2.50	2.500	2.510	0.010	10892	0.0 %	10892					0
5.00	5.000	4.982	-0.018	21593	0.0 %	21593					1
7.50	7.500	7.428	-0.072	32180	0.0 %	32180					1
10.0	10.000	10.059	0.059	43565	0.0 %	43565					1

LB108717
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	85	-		0 ISM02.4	PPB	4/15/2020 11:55	Std
	0.05 S0.05	279	-		0.05 ISM02.4	PPB	4/15/2020 11:57	Std
	0.2 S0.2	899	-		0.2 ISM02.4	PPB	4/15/2020 12:00	Std
	2.5 S2.5	10892	-		2.5 ISM02.4	PPB	4/15/2020 12:02	Std
	5 S5.0	21593	-		5 ISM02.4	PPB	4/15/2020 12:04	Std
	7.5 S7.5	32180	-		7.5 ISM02.4	PPB	4/15/2020 12:07	Std
	10 S10.0	43565	-		10 ISM02.4	PPB	4/15/2020 12:11	Std
ICV	ICV	16615	3.832 -		ISM02.4	PPB	4/15/2020 12:14	SMPL
ICB	ICB	-65	-0.0221 -		ISM02.4	PPB	4/15/2020 12:16	SMPL
CCV016	CCV016	21645	4.9942 -		ISM02.4	PPB	4/15/2020 12:19	SMPL
CCB016	CCB016	-91	-0.0281 -		ISM02.4	PPB	4/15/2020 12:21	SMPL
PB128209BL	PBW209	103	0.0167 -		ISM02.4	PPB	4/15/2020 12:23	SMPL
L2218-01	METC46	373	0.0791 -		ISM02.4	PPB	4/15/2020 12:25	SMPL
L2218-02	METC47	61	0.007 -		ISM02.4	PPB	4/15/2020 12:28	SMPL
L2218-03	METC48	138	0.0248 -		ISM02.4	PPB	4/15/2020 12:30	SMPL
L2218-04	METC48D	119	0.0204 -		ISM02.4	PPB	4/15/2020 12:32	SMPL
L2218-05	METC48S	4541	1.0421 -		ISM02.4	PPB	4/15/2020 12:34	SMPL
CCV017	CCV017	21171	4.8847 -		ISM02.4	PPB	4/15/2020 12:37	SMPL
CCB017	CCB017	-90	-0.0279 -		ISM02.4	PPB	4/15/2020 12:39	SMPL